

### CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1383 SAS No.: Q1383 SDG NO.: Q1383

Continuing Calib Date: 02/24/2025 Initial Calibration Date(s): 02/21/2025 02/21/2025

Continuing Calib Time: 09:50 Initial Calibration Time(s): 19:56 21:32

GC Column: RTX-CLP ID: 0.32 (mm)

| COMPOUND         | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |       | DIFF<br>RT |
|------------------|------------|-----------|------------------------|-------|------------|
| DICAMBA          | 7.34       | 7.34      | 7.24                   | 7.44  | 0.00       |
| 2,4-DCAA         | 7.16       | 7.16      | 7.06                   | 7.26  | 0.00       |
| Dalapon          | 2.60       | 2.60      | 2.50                   | 2.70  | 0.00       |
| DICHLORPROP      | 8.04       | 8.04      | 7.94                   | 8.14  | 0.00       |
| 2,4-D            | 8.27       | 8.27      | 8.17                   | 8.37  | 0.00       |
| 2,4,5-TP(Silvex) | 9.14       | 9.14      | 9.04                   | 9.24  | 0.00       |
| 2,4,5-T          | 9.43       | 9.43      | 9.33                   | 9.53  | 0.00       |
| 2,4-DB           | 10.00      | 10.00     | 9.90                   | 10.10 | 0.00       |
| Dinoseb          | 11.20      | 11.19     | 11.09                  | 11.29 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1383 SAS No.: Q1383 SDG NO.: Q1383

Continuing Calib Date: 02/24/2025 Initial Calibration Date(s): 02/21/2025 02/21/2025

Continuing Calib Time: 09:50 Initial Calibration Time(s): 19:56 21:32

GC Column: RTX-CLP2 ID: 0.32 (mm)

| COMPOUND         | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |       | DIFF<br>RT |
|------------------|------------|-----------|------------------------|-------|------------|
| DICAMBA          | 7.83       | 7.83      | 7.73                   | 7.93  | 0.00       |
| 2,4-DCAA         | 7.63       | 7.63      | 7.53                   | 7.73  | 0.00       |
| Dalapon          | 2.65       | 2.65      | 2.55                   | 2.75  | 0.00       |
| DICHLORPROP      | 8.54       | 8.54      | 8.44                   | 8.64  | 0.00       |
| 2,4-D            | 8.86       | 8.86      | 8.76                   | 8.96  | 0.00       |
| 2,4,5-TP(Silvex) | 9.76       | 9.76      | 9.66                   | 9.86  | 0.00       |
| 2,4,5-T          | 10.17      | 10.17     | 10.07                  | 10.27 | 0.00       |
| 2,4-DB           | 10.73      | 10.73     | 10.63                  | 10.83 | 0.00       |
| Dinoseb          | 11.11      | 11.11     | 11.01                  | 11.21 | 0.00       |