

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1383 SAS No.: Q1383 SDG NO.: Q1383

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 02/21/2025 02/21/2025

Client Sample No.: CCAL01 Date Analyzed: 02/21/2025

Lab Sample No.: HSTDCCC750 Data File : PS029241.D Time Analyzed: 22:44

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.428	9.329	9.529	695.300	712.500	-2.4
2,4,5-TP(Silvex)	9.137	9.038	9.238	693.810	712.500	-2.6
2,4-D	8.271	8.171	8.371	677.640	705.000	-3.9
2,4-DB	9.997	9.898	10.098	702.860	712.500	-1.4
2,4-DCAA	7.159	7.060	7.260	720.300	750.000	-4.0
Dalapon	2.599	2.500	2.700	658.620	682.500	-3.5
DICAMBA	7.342	7.243	7.443	690.300	705.000	-2.1
DICHLORPROP	8.042	7.942	8.142	670.150	705.000	-4.9
Dinoseb	11.194	11.094	11.294	687.340	705.000	-2.5

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1383 SAS No.: Q1383 SDG NO.: Q1383

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/21/2025 02/21/2025

Client Sample No.: CCAL01 Date Analyzed: 02/21/2025

Lab Sample No.: HSTDCCC750 Data File : PS029241.D Time Analyzed: 22:44

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.170	10.070	10.270	707.200	712.500	-0.7
2,4,5-TP(Silvex)	9.755	9.656	9.856	708.970	712.500	-0.5
2,4-D	8.861	8.761	8.961	687.930	705.000	-2.4
2,4-DB	10.732	10.634	10.834	690.110	712.500	-3.1
2,4-DCAA	7.631	7.532	7.732	730.910	750.000	-2.5
Dalapon	2.647	2.551	2.751	660.620	682.500	-3.2
DICAMBA	7.826	7.726	7.926	700.940	705.000	-0.6
DICHLORPROP	8.535	8.436	8.636	683.170	705.000	-3.1
Dinoseb	11.109	11.009	11.209	698.050	705.000	-1.0