

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1383 SAS No.: Q1383 SDG NO.: Q1383

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 02/21/2025 02/21/2025

Client Sample No.: CCAL02 Date Analyzed: 02/22/2025

Lab Sample No.: HSTDCCC750 Data File : PS029252.D Time Analyzed: 03:08

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.428	9.329	9.529	705.420	712.500	-1.0
2,4,5-TP(Silvex)	9.137	9.038	9.238	704.100	712.500	-1.2
2,4-D	8.270	8.171	8.371	689.170	705.000	-2.2
2,4-DB	9.997	9.898	10.098	705.390	712.500	-1.0
2,4-DCAA	7.159	7.060	7.260	732.950	750.000	-2.3
Dalapon	2.598	2.500	2.700	659.270	682.500	-3.4
DICAMBA	7.342	7.243	7.443	700.100	705.000	-0.7
DICHLORPROP	8.041	7.942	8.142	681.040	705.000	-3.4
Dinoseb	11.193	11.094	11.294	695.210	705.000	-1.4

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1383 SAS No.: Q1383 SDG NO.: Q1383

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/21/2025 02/21/2025

Client Sample No.: CCAL02 Date Analyzed: 02/22/2025

Lab Sample No.: HSTDCCC750 Data File : PS029252.D Time Analyzed: 03:08

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.169	10.070	10.270	720.010	712.500	1.1
2,4,5-TP(Silvex)	9.754	9.656	9.856	720.050	712.500	1.1
2,4-D	8.860	8.761	8.961	703.860	705.000	-0.2
2,4-DB	10.732	10.634	10.834	702.690	712.500	-1.4
2,4-DCAA	7.630	7.532	7.732	744.150	750.000	-0.8
Dalapon	2.649	2.551	2.751	660.090	682.500	-3.3
DICAMBA	7.825	7.726	7.926	712.870	705.000	1.1
DICHLORPROP	8.535	8.436	8.636	700.090	705.000	-0.7
Dinoseb	11.108	11.009	11.209	705.230	705.000	0.0