

### CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1383 SAS No.: Q1383 SDG NO.: Q1383

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 02/21/2025 02/21/2025

Client Sample No.: CCAL03 Date Analyzed: 02/24/2025

Lab Sample No.: HSTDCCC750 Data File : PS029255.D Time Analyzed: 09:50

| COMPOUND         | RT     | RT WINDOW<br>FROM TO |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D  |
|------------------|--------|----------------------|--------|--------------------|-------------------|-----|
| 2,4,5-T          | 9.430  | 9.329                | 9.529  | 732.010            | 712.500           | 2.7 |
| 2,4,5-TP(Silvex) | 9.139  | 9.038                | 9.238  | 735.940            | 712.500           | 3.3 |
| 2,4-D            | 8.273  | 8.171                | 8.371  | 713.790            | 705.000           | 1.2 |
| 2,4-DB           | 9.998  | 9.898                | 10.098 | 734.640            | 712.500           | 3.1 |
| 2,4-DCAA         | 7.160  | 7.060                | 7.260  | 759.100            | 750.000           | 1.2 |
| Dalapon          | 2.602  | 2.500                | 2.700  | 701.670            | 682.500           | 2.8 |
| DICAMBA          | 7.343  | 7.243                | 7.443  | 726.940            | 705.000           | 3.1 |
| DICHLORPROP      | 8.043  | 7.942                | 8.142  | 705.700            | 705.000           | 0.1 |
| Dinoseb          | 11.195 | 11.094               | 11.294 | 721.410            | 705.000           | 2.3 |

### CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1383 SAS No.: Q1383 SDG NO.: Q1383

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/21/2025 02/21/2025

Client Sample No.: CCAL03 Date Analyzed: 02/24/2025

Lab Sample No.: HSTDCCC750 Data File : PS029255.D Time Analyzed: 09:50

| COMPOUND         | RT     | RT WINDOW<br>FROM TO |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|------------------|--------|----------------------|--------|--------------------|-------------------|------|
| 2,4,5-T          | 10.171 | 10.070               | 10.270 | 720.880            | 712.500           | 1.2  |
| 2,4,5-TP(Silvex) | 9.756  | 9.656                | 9.856  | 722.380            | 712.500           | 1.4  |
| 2,4-D            | 8.861  | 8.761                | 8.961  | 697.740            | 705.000           | -1.0 |
| 2,4-DB           | 10.733 | 10.634               | 10.834 | 702.440            | 712.500           | -1.4 |
| 2,4-DCAA         | 7.632  | 7.532                | 7.732  | 740.920            | 750.000           | -1.2 |
| Dalapon          | 2.651  | 2.551                | 2.751  | 671.610            | 682.500           | -1.6 |
| DICAMBA          | 7.826  | 7.726                | 7.926  | 712.330            | 705.000           | 1.0  |
| DICHLORPROP      | 8.536  | 8.436                | 8.636  | 693.390            | 705.000           | -1.6 |
| Dinoseb          | 11.109 | 11.009               | 11.209 | 708.660            | 705.000           | 0.5  |