

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1383 SAS No.: Q1383 SDG NO.: Q1383

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 02/21/2025 02/21/2025

Client Sample No.: CCAL04 Date Analyzed: 02/24/2025

Lab Sample No.: HSTDCCC750 Data File : PS029267.D Time Analyzed: 15:59

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.429	9.329	9.529	749.180	712.500	5.1
2,4,5-TP(Silvex)	9.139	9.038	9.238	750.490	712.500	5.3
2,4-D	8.271	8.171	8.371	730.290	705.000	3.6
2,4-DB	9.999	9.898	10.098	752.770	712.500	5.7
2,4-DCAA	7.160	7.060	7.260	780.130	750.000	4.0
Dalapon	2.600	2.500	2.700	714.160	682.500	4.6
DICAMBA	7.343	7.243	7.443	741.780	705.000	5.2
DICHLORPROP	8.043	7.942	8.142	723.240	705.000	2.6
Dinoseb	11.194	11.094	11.294	753.630	705.000	6.9

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1383 SAS No.: Q1383 SDG NO.: Q1383

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/21/2025 02/21/2025

Client Sample No.: CCAL04 Date Analyzed: 02/24/2025

Lab Sample No.: HSTDCCC750 Data File : PS029267.D Time Analyzed: 15:59

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.169	10.070	10.270	742.180	712.500	4.2
2,4,5-TP(Silvex)	9.755	9.656	9.856	741.540	712.500	4.1
2,4-D	8.860	8.761	8.961	719.270	705.000	2.0
2,4-DB	10.732	10.634	10.834	714.910	712.500	0.3
2,4-DCAA	7.631	7.532	7.732	758.850	750.000	1.2
Dalapon	2.649	2.551	2.751	690.150	682.500	1.1
DICAMBA	7.825	7.726	7.926	732.490	705.000	3.9
DICHLORPROP	8.534	8.436	8.636	713.470	705.000	1.2
Dinoseb	11.108	11.009	11.209	727.580	705.000	3.2