

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1383 SAS No.: Q1383 SDG NO.: Q1383

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 02/21/2025 02/21/2025

Client Sample No.: CCAL06 Date Analyzed: 02/26/2025

Lab Sample No.: HSTDCCC750 Data File : PS029303.D Time Analyzed: 19:23

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.427	9.329	9.529	765.500	712.500	7.4
2,4,5-TP(Silvex)	9.137	9.038	9.238	766.410	712.500	7.6
2,4-D	8.269	8.171	8.371	750.970	705.000	6.5
2,4-DB	9.996	9.898	10.098	768.040	712.500	7.8
2,4-DCAA	7.158	7.060	7.260	796.740	750.000	6.2
Dalapon	2.600	2.500	2.700	689.240	682.500	1.0
DICAMBA	7.341	7.243	7.443	761.360	705.000	8.0
DICHLORPROP	8.040	7.942	8.142	740.310	705.000	5.0
Dinoseb	11.193	11.094	11.294	747.850	705.000	6.1

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1383 SAS No.: Q1383 SDG NO.: Q1383

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/21/2025 02/21/2025

Client Sample No.: CCAL06 Date Analyzed: 02/26/2025

Lab Sample No.: HSTDCCC750 Data File : PS029303.D Time Analyzed: 19:23

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.165	10.070	10.270	764.870	712.500	7.4
2,4,5-TP(Silvex)	9.750	9.656	9.856	764.830	712.500	7.3
2,4-D	8.856	8.761	8.961	743.990	705.000	5.5
2,4-DB	10.728	10.634	10.834	714.720	712.500	0.3
2,4-DCAA	7.627	7.532	7.732	778.000	750.000	3.7
Dalapon	2.647	2.551	2.751	667.410	682.500	-2.2
DICAMBA	7.821	7.726	7.926	753.200	705.000	6.8
DICHLORPROP	8.531	8.436	8.636	740.930	705.000	5.1
Dinoseb	11.103	11.009	11.209	736.630	705.000	4.5