

### CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1383 SAS No.: Q1383 SDG NO.: Q1383

Continuing Calib Date: 02/20/2025 Initial Calibration Date(s): 01/28/2025 01/28/2025

Continuing Calib Time: 10:38 Initial Calibration Time(s): 09:37 17:30

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|-------|------------|
|                      |     |            |           | FROM      | TO    |            |
| Aroclor-1016-1       | (1) | 5.68       | 5.69      | 5.59      | 5.79  | 0.01       |
| Aroclor-1016-2       | (2) | 5.70       | 5.71      | 5.61      | 5.81  | 0.01       |
| Aroclor-1016-3       | (3) | 5.76       | 5.78      | 5.68      | 5.88  | 0.02       |
| Aroclor-1016-4       | (4) | 5.86       | 5.87      | 5.77      | 5.97  | 0.01       |
| Aroclor-1016-5       | (5) | 6.15       | 6.17      | 6.07      | 6.27  | 0.02       |
| Aroclor-1260-1       | (1) | 7.27       | 7.29      | 7.19      | 7.39  | 0.02       |
| Aroclor-1260-2       | (2) | 7.53       | 7.54      | 7.44      | 7.64  | 0.01       |
| Aroclor-1260-3       | (3) | 7.89       | 7.90      | 7.80      | 8.00  | 0.01       |
| Aroclor-1260-4       | (4) | 8.11       | 8.13      | 8.03      | 8.23  | 0.02       |
| Aroclor-1260-5       | (5) | 8.43       | 8.45      | 8.35      | 8.55  | 0.02       |
| Tetrachloro-m-xylene |     | 4.52       | 4.54      | 4.44      | 4.64  | 0.02       |
| Decachlorobiphenyl   |     | 10.25      | 10.28     | 10.18     | 10.38 | 0.03       |

### CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1383 SAS No.: Q1383 SDG NO.: Q1383

Continuing Calib Date: 02/20/2025 Initial Calibration Date(s): 01/28/2025 01/28/2025

Continuing Calib Time: 10:38 Initial Calibration Time(s): 09:37 17:30

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1016-1       | (1) | 4.92       | 4.93      | 4.83      | 5.03 | 0.01       |
| Aroclor-1016-2       | (2) | 4.94       | 4.95      | 4.85      | 5.05 | 0.01       |
| Aroclor-1016-3       | (3) | 5.12       | 5.13      | 5.03      | 5.23 | 0.01       |
| Aroclor-1016-4       | (4) | 5.16       | 5.17      | 5.07      | 5.27 | 0.01       |
| Aroclor-1016-5       | (5) | 5.37       | 5.39      | 5.29      | 5.49 | 0.02       |
| Aroclor-1260-1       | (1) | 6.41       | 6.43      | 6.33      | 6.53 | 0.02       |
| Aroclor-1260-2       | (2) | 6.60       | 6.62      | 6.52      | 6.72 | 0.02       |
| Aroclor-1260-3       | (3) | 6.75       | 6.77      | 6.67      | 6.87 | 0.02       |
| Aroclor-1260-4       | (4) | 7.23       | 7.24      | 7.14      | 7.34 | 0.02       |
| Aroclor-1260-5       | (5) | 7.47       | 7.48      | 7.38      | 7.58 | 0.01       |
| Tetrachloro-m-xylene |     | 3.83       | 3.84      | 3.74      | 3.94 | 0.01       |
| Decachlorobiphenyl   |     | 8.89       | 8.91      | 8.81      | 9.01 | 0.02       |