

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1383 SAS No.: Q1383 SDG NO.: Q1383

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL02 Date Analyzed: 02/19/2025

Lab Sample No.: AR1254CCC500 Data File : PP069860.D Time Analyzed: 21:23

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1254-1	6.525	6.439	6.639	464.010	500.000	-7.2
Aroclor-1254-2	6.742	6.656	6.856	546.680	500.000	9.3
Aroclor-1254-3	7.105	7.019	7.219	516.820	500.000	3.4
Aroclor-1254-4	7.387	7.301	7.501	526.820	500.000	5.4
Aroclor-1254-5	7.803	7.719	7.919	478.070	500.000	-4.4
Decachlorobiphenyl	10.250	10.175	10.375	51.860	50.000	3.7
Tetrachloro-m-xylene	4.524	4.434	4.634	54.550	50.000	9.1

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1383 SAS No.: Q1383 SDG NO.: Q1383

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL02 Date Analyzed: 02/19/2025

Lab Sample No.: AR1254CCC500 Data File : PP069860.D Time Analyzed: 21:23

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1254-1	5.727	5.642	5.842	504.880	500.000	1.0
Aroclor-1254-2	5.875	5.790	5.990	499.940	500.000	0.0
Aroclor-1254-3	6.280	6.195	6.395	500.560	500.000	0.1
Aroclor-1254-4	6.509	6.423	6.623	507.630	500.000	1.5
Aroclor-1254-5	6.928	6.844	7.044	497.110	500.000	-0.6
Decachlorobiphenyl	8.888	8.809	9.009	50.310	50.000	0.6
Tetrachloro-m-xylene	3.830	3.740	3.940	53.290	50.000	6.6