

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1383 SAS No.: Q1383 SDG NO.: Q1383

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL03 Date Analyzed: 02/20/2025

Lab Sample No.: AR1660CCC500 Data File : PP069872.D Time Analyzed: 01:26

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.678	5.592	5.792	497.890	500.000	-0.4
Aroclor-1016-2	5.701	5.614	5.814	513.100	500.000	2.6
Aroclor-1016-3	5.763	5.676	5.876	501.620	500.000	0.3
Aroclor-1016-4	5.860	5.774	5.974	510.080	500.000	2.0
Aroclor-1016-5	6.153	6.067	6.267	466.370	500.000	-6.7
Aroclor-1260-1	7.272	7.187	7.387	522.330	500.000	4.5
Aroclor-1260-2	7.527	7.441	7.641	496.550	500.000	-0.7
Aroclor-1260-3	7.885	7.800	8.000	511.610	500.000	2.3
Aroclor-1260-4	8.109	8.025	8.225	488.860	500.000	-2.2
Aroclor-1260-5	8.431	8.347	8.547	497.190	500.000	-0.6
Decachlorobiphenyl	10.254	10.179	10.379	51.550	50.000	3.1
Tetrachloro-m-xylene	4.525	4.438	4.638	53.110	50.000	6.2

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1383 SAS No.: Q1383 SDG NO.: Q1383

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL03 Date Analyzed: 02/20/2025

Lab Sample No.: AR1660CCC500 Data File : PP069872.D Time Analyzed: 01:26

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.920	4.834	5.034	451.250	500.000	-9.8
Aroclor-1016-2	4.940	4.853	5.053	471.590	500.000	-5.7
Aroclor-1016-3	5.117	5.030	5.230	481.160	500.000	-3.8
Aroclor-1016-4	5.159	5.072	5.272	446.290	500.000	-10.7
Aroclor-1016-5	5.373	5.288	5.488	471.380	500.000	-5.7
Aroclor-1260-1	6.411	6.326	6.526	543.110	500.000	8.6
Aroclor-1260-2	6.599	6.515	6.715	543.360	500.000	8.7
Aroclor-1260-3	6.754	6.670	6.870	471.020	500.000	-5.8
Aroclor-1260-4	7.226	7.142	7.342	514.580	500.000	2.9
Aroclor-1260-5	7.466	7.384	7.584	537.280	500.000	7.5
Decachlorobiphenyl	8.889	8.811	9.011	49.940	50.000	-0.1
Tetrachloro-m-xylene	3.830	3.740	3.940	49.520	50.000	-1.0