

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1383 SAS No.: Q1383 SDG NO.: Q1383

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL04 Date Analyzed: 02/20/2025

Lab Sample No.: AR1254CCC500 Data File : PP069875.D Time Analyzed: 02:31

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1254-1	6.530	6.439	6.639	516.830	500.000	3.4
Aroclor-1254-2	6.746	6.656	6.856	581.880	500.000	16.4
Aroclor-1254-3	7.108	7.019	7.219	560.730	500.000	12.1
Aroclor-1254-4	7.390	7.301	7.501	557.460	500.000	11.5
Aroclor-1254-5	7.808	7.719	7.919	528.510	500.000	5.7
Decachlorobiphenyl	10.254	10.175	10.375	53.220	50.000	6.4
Tetrachloro-m-xylene	4.526	4.434	4.634	55.750	50.000	11.5

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1383 SAS No.: Q1383 SDG NO.: Q1383

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL04 Date Analyzed: 02/20/2025

Lab Sample No.: AR1254CCC500 Data File : PP069875.D Time Analyzed: 02:31

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1254-1	5.726	5.642	5.842	529.150	500.000	5.8
Aroclor-1254-2	5.875	5.790	5.990	533.990	500.000	6.8
Aroclor-1254-3	6.280	6.195	6.395	522.840	500.000	4.6
Aroclor-1254-4	6.508	6.423	6.623	529.870	500.000	6.0
Aroclor-1254-5	6.927	6.844	7.044	518.160	500.000	3.6
Decachlorobiphenyl	8.889	8.809	9.009	48.300	50.000	-3.4
Tetrachloro-m-xylene	3.829	3.740	3.940	53.960	50.000	7.9