

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1383 SAS No.: Q1383 SDG NO.: Q1383

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL05 Date Analyzed: 02/20/2025

Lab Sample No.: AR1660CCC500 Data File : PP069885.D Time Analyzed: 06:02

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.675	5.592	5.792	559.400	500.000	11.9
Aroclor-1016-2	5.698	5.614	5.814	520.300	500.000	4.1
Aroclor-1016-3	5.761	5.676	5.876	500.940	500.000	0.2
Aroclor-1016-4	5.858	5.774	5.974	511.830	500.000	2.4
Aroclor-1016-5	6.150	6.067	6.267	509.960	500.000	2.0
Aroclor-1260-1	7.271	7.187	7.387	549.670	500.000	9.9
Aroclor-1260-2	7.524	7.441	7.641	516.550	500.000	3.3
Aroclor-1260-3	7.884	7.800	8.000	532.790	500.000	6.6
Aroclor-1260-4	8.108	8.025	8.225	508.370	500.000	1.7
Aroclor-1260-5	8.428	8.347	8.547	507.550	500.000	1.5
Decachlorobiphenyl	10.252	10.179	10.379	51.700	50.000	3.4
Tetrachloro-m-xylene	4.522	4.438	4.638	54.320	50.000	8.6

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1383 SAS No.: Q1383 SDG NO.: Q1383

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL05 Date Analyzed: 02/20/2025

Lab Sample No.: AR1660CCC500 Data File : PP069885.D Time Analyzed: 06:02

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.920	4.834	5.034	451.840	500.000	-9.6
Aroclor-1016-2	4.940	4.853	5.053	478.020	500.000	-4.4
Aroclor-1016-3	5.117	5.030	5.230	493.950	500.000	-1.2
Aroclor-1016-4	5.159	5.072	5.272	448.920	500.000	-10.2
Aroclor-1016-5	5.374	5.288	5.488	492.010	500.000	-1.6
Aroclor-1260-1	6.412	6.326	6.526	527.160	500.000	5.4
Aroclor-1260-2	6.600	6.515	6.715	515.170	500.000	3.0
Aroclor-1260-3	6.754	6.670	6.870	455.350	500.000	-8.9
Aroclor-1260-4	7.227	7.142	7.342	507.470	500.000	1.5
Aroclor-1260-5	7.468	7.384	7.584	530.560	500.000	6.1
Decachlorobiphenyl	8.890	8.811	9.011	50.910	50.000	1.8
Tetrachloro-m-xylene	3.830	3.740	3.940	51.220	50.000	2.4