

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1383 SAS No.: Q1383 SDG NO.: Q1383

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL06 Date Analyzed: 02/20/2025

Lab Sample No.: AR1660CCC500 Data File : PP069932.D Time Analyzed: 21:28

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.676	5.592	5.792	502.680	500.000	0.5
Aroclor-1016-2	5.700	5.614	5.814	536.020	500.000	7.2
Aroclor-1016-3	5.763	5.676	5.876	518.730	500.000	3.7
Aroclor-1016-4	5.860	5.774	5.974	530.710	500.000	6.1
Aroclor-1016-5	6.153	6.067	6.267	488.960	500.000	-2.2
Aroclor-1260-1	7.272	7.187	7.387	556.270	500.000	11.3
Aroclor-1260-2	7.526	7.441	7.641	524.280	500.000	4.9
Aroclor-1260-3	7.884	7.800	8.000	533.350	500.000	6.7
Aroclor-1260-4	8.109	8.025	8.225	501.940	500.000	0.4
Aroclor-1260-5	8.429	8.347	8.547	514.160	500.000	2.8
Decachlorobiphenyl	10.252	10.179	10.379	53.380	50.000	6.8
Tetrachloro-m-xylene	4.525	4.438	4.638	53.680	50.000	7.4

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1383 SAS No.: Q1383 SDG NO.: Q1383

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL06 Date Analyzed: 02/20/2025

Lab Sample No.: AR1660CCC500 Data File : PP069932.D Time Analyzed: 21:28

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.919	4.834	5.034	488.890	500.000	-2.2
Aroclor-1016-2	4.938	4.853	5.053	492.730	500.000	-1.5
Aroclor-1016-3	5.116	5.030	5.230	503.840	500.000	0.8
Aroclor-1016-4	5.158	5.072	5.272	461.780	500.000	-7.6
Aroclor-1016-5	5.373	5.288	5.488	583.650	500.000	16.7
Aroclor-1260-1	6.410	6.326	6.526	543.080	500.000	8.6
Aroclor-1260-2	6.599	6.515	6.715	544.190	500.000	8.8
Aroclor-1260-3	6.752	6.670	6.870	470.720	500.000	-5.9
Aroclor-1260-4	7.225	7.142	7.342	526.380	500.000	5.3
Aroclor-1260-5	7.466	7.384	7.584	537.060	500.000	7.4
Decachlorobiphenyl	8.887	8.811	9.011	49.230	50.000	-1.5
Tetrachloro-m-xylene	3.829	3.740	3.940	51.980	50.000	4.0