

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1383 SAS No.: Q1383 SDG NO.: Q1383

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL08 Date Analyzed: 02/21/2025

Lab Sample No.: AR1660CCC500 Data File : PP069947.D Time Analyzed: 01:53

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.676	5.592	5.792	564.420	500.000	12.9
Aroclor-1016-2	5.699	5.614	5.814	551.540	500.000	10.3
Aroclor-1016-3	5.762	5.676	5.876	535.310	500.000	7.1
Aroclor-1016-4	5.859	5.774	5.974	550.910	500.000	10.2
Aroclor-1016-5	6.151	6.067	6.267	514.090	500.000	2.8
Aroclor-1260-1	7.269	7.187	7.387	556.970	500.000	11.4
Aroclor-1260-2	7.525	7.441	7.641	548.930	500.000	9.8
Aroclor-1260-3	7.883	7.800	8.000	557.900	500.000	11.6
Aroclor-1260-4	8.108	8.025	8.225	518.950	500.000	3.8
Aroclor-1260-5	8.428	8.347	8.547	527.430	500.000	5.5
Decachlorobiphenyl	10.250	10.179	10.379	54.850	50.000	9.7
Tetrachloro-m-xylene	4.523	4.438	4.638	56.870	50.000	13.7

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1383 SAS No.: Q1383 SDG NO.: Q1383

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL08 Date Analyzed: 02/21/2025

Lab Sample No.: AR1660CCC500 Data File : PP069947.D Time Analyzed: 01:53

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.918	4.834	5.034	503.540	500.000	0.7
Aroclor-1016-2	4.938	4.853	5.053	481.370	500.000	-3.7
Aroclor-1016-3	5.115	5.030	5.230	518.770	500.000	3.8
Aroclor-1016-4	5.157	5.072	5.272	470.280	500.000	-5.9
Aroclor-1016-5	5.371	5.288	5.488	526.850	500.000	5.4
Aroclor-1260-1	6.410	6.326	6.526	545.470	500.000	9.1
Aroclor-1260-2	6.598	6.515	6.715	545.400	500.000	9.1
Aroclor-1260-3	6.752	6.670	6.870	485.550	500.000	-2.9
Aroclor-1260-4	7.225	7.142	7.342	540.170	500.000	8.0
Aroclor-1260-5	7.465	7.384	7.584	554.170	500.000	10.8
Decachlorobiphenyl	8.886	8.811	9.011	50.260	50.000	0.5
Tetrachloro-m-xylene	3.829	3.740	3.940	54.320	50.000	8.6