

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1383 SAS No.: Q1383 SDG NO.: Q1383

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL09 Date Analyzed: 02/21/2025

Lab Sample No.: AR1254CCC500 Data File : PP069950.D Time Analyzed: 02:42

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1254-1	6.524	6.439	6.639	489.510	500.000	-2.1
Aroclor-1254-2	6.742	6.656	6.856	551.740	500.000	10.3
Aroclor-1254-3	7.105	7.019	7.219	519.320	500.000	3.9
Aroclor-1254-4	7.387	7.301	7.501	526.750	500.000	5.4
Aroclor-1254-5	7.803	7.719	7.919	517.250	500.000	3.5
Decachlorobiphenyl	10.250	10.175	10.375	53.640	50.000	7.3
Tetrachloro-m-xylene	4.523	4.434	4.634	55.550	50.000	11.1

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1383 SAS No.: Q1383 SDG NO.: Q1383

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL09 Date Analyzed: 02/21/2025

Lab Sample No.: AR1254CCC500 Data File : PP069950.D Time Analyzed: 02:42

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1254-1	5.725	5.642	5.842	511.430	500.000	2.3
Aroclor-1254-2	5.874	5.790	5.990	516.810	500.000	3.4
Aroclor-1254-3	6.279	6.195	6.395	513.230	500.000	2.6
Aroclor-1254-4	6.507	6.423	6.623	523.310	500.000	4.7
Aroclor-1254-5	6.927	6.844	7.044	528.510	500.000	5.7
Decachlorobiphenyl	8.887	8.809	9.009	51.720	50.000	3.4
Tetrachloro-m-xylene	3.829	3.740	3.940	50.680	50.000	1.4