

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1383 SAS No.: Q1383 SDG NO.: Q1383

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL10 Date Analyzed: 02/21/2025

Lab Sample No.: AR1660CCC500 Data File : PP069953.D Time Analyzed: 09:16

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.678	5.592	5.792	600.590	500.000	20.1
Aroclor-1016-2	5.702	5.614	5.814	532.170	500.000	6.4
Aroclor-1016-3	5.764	5.676	5.876	520.050	500.000	4.0
Aroclor-1016-4	5.861	5.774	5.974	532.460	500.000	6.5
Aroclor-1016-5	6.154	6.067	6.267	494.630	500.000	-1.1
Aroclor-1260-1	7.274	7.187	7.387	512.450	500.000	2.5
Aroclor-1260-2	7.527	7.441	7.641	508.810	500.000	1.8
Aroclor-1260-3	7.885	7.800	8.000	521.910	500.000	4.4
Aroclor-1260-4	8.110	8.025	8.225	494.890	500.000	-1.0
Aroclor-1260-5	8.432	8.347	8.547	513.580	500.000	2.7
Decachlorobiphenyl	10.257	10.179	10.379	54.010	50.000	8.0
Tetrachloro-m-xylene	4.526	4.438	4.638	55.440	50.000	10.9

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1383 SAS No.: Q1383 SDG NO.: Q1383

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL10 Date Analyzed: 02/21/2025

Lab Sample No.: AR1660CCC500 Data File : PP069953.D Time Analyzed: 09:16

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.921	4.834	5.034	511.830	500.000	2.4
Aroclor-1016-2	4.940	4.853	5.053	488.550	500.000	-2.3
Aroclor-1016-3	5.119	5.030	5.230	510.960	500.000	2.2
Aroclor-1016-4	5.160	5.072	5.272	510.240	500.000	2.0
Aroclor-1016-5	5.375	5.288	5.488	493.980	500.000	-1.2
Aroclor-1260-1	6.413	6.326	6.526	508.430	500.000	1.7
Aroclor-1260-2	6.601	6.515	6.715	526.150	500.000	5.2
Aroclor-1260-3	6.755	6.670	6.870	462.750	500.000	-7.5
Aroclor-1260-4	7.227	7.142	7.342	528.470	500.000	5.7
Aroclor-1260-5	7.468	7.384	7.584	543.670	500.000	8.7
Decachlorobiphenyl	8.889	8.811	9.011	50.500	50.000	1.0
Tetrachloro-m-xylene	3.832	3.740	3.940	52.870	50.000	5.7