

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1383 SAS No.: Q1383 SDG NO.: Q1383

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL11 Date Analyzed: 02/21/2025

Lab Sample No.: AR1254CCC500 Data File : PP069956.D Time Analyzed: 10:04

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1254-1	6.524	6.439	6.639	445.040	500.000	-11.0
Aroclor-1254-2	6.742	6.656	6.856	521.560	500.000	4.3
Aroclor-1254-3	7.105	7.019	7.219	494.310	500.000	-1.1
Aroclor-1254-4	7.387	7.301	7.501	511.750	500.000	2.4
Aroclor-1254-5	7.803	7.719	7.919	494.290	500.000	-1.1
Decachlorobiphenyl	10.249	10.175	10.375	51.620	50.000	3.2
Tetrachloro-m-xylene	4.523	4.434	4.634	53.040	50.000	6.1

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1383 SAS No.: Q1383 SDG NO.: Q1383

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL11 Date Analyzed: 02/21/2025

Lab Sample No.: AR1254CCC500 Data File : PP069956.D Time Analyzed: 10:04

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1254-1	5.726	5.642	5.842	481.880	500.000	-3.6
Aroclor-1254-2	5.874	5.790	5.990	491.440	500.000	-1.7
Aroclor-1254-3	6.279	6.195	6.395	489.360	500.000	-2.1
Aroclor-1254-4	6.507	6.423	6.623	485.970	500.000	-2.8
Aroclor-1254-5	6.927	6.844	7.044	485.750	500.000	-2.9
Decachlorobiphenyl	8.886	8.809	9.009	47.050	50.000	-5.9
Tetrachloro-m-xylene	3.829	3.740	3.940	52.180	50.000	4.4