

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1383 SAS No.: Q1383 SDG NO.: Q1383

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL12 Date Analyzed: 02/21/2025

Lab Sample No.: AR1660CCC500 Data File : PP069967.D Time Analyzed: 14:58

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.677	5.592	5.792	522.230	500.000	4.4
Aroclor-1016-2	5.701	5.614	5.814	459.800	500.000	-8.0
Aroclor-1016-3	5.763	5.676	5.876	437.360	500.000	-12.5
Aroclor-1016-4	5.861	5.774	5.974	452.610	500.000	-9.5
Aroclor-1016-5	6.152	6.067	6.267	461.540	500.000	-7.7
Aroclor-1260-1	7.272	7.187	7.387	436.560	500.000	-12.7
Aroclor-1260-2	7.526	7.441	7.641	434.730	500.000	-13.1
Aroclor-1260-3	7.884	7.800	8.000	447.160	500.000	-10.6
Aroclor-1260-4	8.108	8.025	8.225	412.700	500.000	-17.5
Aroclor-1260-5	8.429	8.347	8.547	455.700	500.000	-8.9
Decachlorobiphenyl	10.251	10.179	10.379	47.880	50.000	-4.2
Tetrachloro-m-xylene	4.525	4.438	4.638	52.560	50.000	5.1

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1383 SAS No.: Q1383 SDG NO.: Q1383

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL12 Date Analyzed: 02/21/2025

Lab Sample No.: AR1660CCC500 Data File : PP069967.D Time Analyzed: 14:58

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.919	4.834	5.034	447.700	500.000	-10.5
Aroclor-1016-2	4.939	4.853	5.053	510.780	500.000	2.2
Aroclor-1016-3	5.116	5.030	5.230	493.140	500.000	-1.4
Aroclor-1016-4	5.157	5.072	5.272	482.780	500.000	-3.4
Aroclor-1016-5	5.373	5.288	5.488	477.600	500.000	-4.5
Aroclor-1260-1	6.410	6.326	6.526	451.140	500.000	-9.8
Aroclor-1260-2	6.598	6.515	6.715	470.090	500.000	-6.0
Aroclor-1260-3	6.752	6.670	6.870	425.860	500.000	-14.8
Aroclor-1260-4	7.225	7.142	7.342	460.000	500.000	-8.0
Aroclor-1260-5	7.465	7.384	7.584	468.000	500.000	-6.4
Decachlorobiphenyl	8.886	8.811	9.011	41.610	50.000	-16.8
Tetrachloro-m-xylene	3.830	3.740	3.940	51.860	50.000	3.7