

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1383 SAS No.: Q1383 SDG NO.: Q1383

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL13 Date Analyzed: 02/21/2025

Lab Sample No.: AR1254CCC500 Data File : PP069970.D Time Analyzed: 15:47

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1254-1	6.524	6.439	6.639	460.970	500.000	-7.8
Aroclor-1254-2	6.742	6.656	6.856	502.620	500.000	0.5
Aroclor-1254-3	7.104	7.019	7.219	520.810	500.000	4.2
Aroclor-1254-4	7.386	7.301	7.501	468.850	500.000	-6.2
Aroclor-1254-5	7.802	7.719	7.919	455.850	500.000	-8.8
Decachlorobiphenyl	10.248	10.175	10.375	48.990	50.000	-2.0
Tetrachloro-m-xylene	4.523	4.434	4.634	52.920	50.000	5.8

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q1383 SAS No.: Q1383 SDG NO.: Q1383

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL13 Date Analyzed: 02/21/2025

Lab Sample No.: AR1254CCC500 Data File : PP069970.D Time Analyzed: 15:47

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1254-1	5.725	5.642	5.842	463.930	500.000	-7.2
Aroclor-1254-2	5.873	5.790	5.990	461.930	500.000	-7.6
Aroclor-1254-3	6.278	6.195	6.395	462.690	500.000	-7.5
Aroclor-1254-4	6.506	6.423	6.623	467.520	500.000	-6.5
Aroclor-1254-5	6.926	6.844	7.044	469.980	500.000	-6.0
Decachlorobiphenyl	8.886	8.809	9.009	43.890	50.000	-12.2
Tetrachloro-m-xylene	3.829	3.740	3.940	52.070	50.000	4.1