

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP012825\
 Data File : PP069291.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Jan 2025 17:14
 Operator : YP\AJ
 Sample : AR1268ICC250
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 28 17:28:51 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012825.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 28 17:28:39 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.553	3.841	39693299	25406430	26.359	26.193
2) SA Decachlor...	10.293	8.912	50826883	48064273	26.937	25.976
Target Compounds						
41) L9 AR-1268-1	8.778	7.768	50705640	40061895	267.938	259.651
42) L9 AR-1268-2	8.872	7.832	44793849	36640221	266.509	257.814
43) L9 AR-1268-3	9.106	8.041	38627120	31135456	268.197	257.001
44) L9 AR-1268-4	9.527	8.336	16972703	14773829	266.811	261.656
45) L9 AR-1268-5	9.947	8.642	109.1E6	95443492	265.618	250.089



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP012825\
Data File : PP069291.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Jan 2025 17:14
Operator : YP\AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 28 17:28:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012825.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 28 17:28:39 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.553 min
Delta R.T.: 0.000 min
Response: 39693299
Conc: 26.36 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.841 min
Delta R.T.: 0.000 min
Response: 25406430
Conc: 26.19 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.293 min
Delta R.T.: 0.000 min
Response: 50826883
Conc: 26.94 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.912 min
Delta R.T.: 0.000 min
Response: 48064273
Conc: 25.98 ng/ml



#41 AR-1268-1


R.T.: 8.778 min
Delta R.T.: 0.000 min
Response: 50705640
Conc: 267.94 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250




#41 AR-1268-1

R.T.: 7.768 min
Delta R.T.: 0.000 min
Response: 40061895
Conc: 259.65 ng/ml




#42 AR-1268-2

R.T.: 8.872 min
Delta R.T.: 0.000 min
Response: 44793849
Conc: 266.51 ng/ml



#42 AR-1268-2


R.T.: 7.832 min
Delta R.T.: 0.000 min
Response: 36640221
Conc: 257.81 ng/ml



#43 AR-1268-3


R.T.: 9.106 min
Delta R.T.: 0.000 min
Response: 38627120
Conc: 268.20 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250




#43 AR-1268-3

R.T.: 8.041 min
Delta R.T.: 0.000 min
Response: 31135456
Conc: 257.00 ng/ml




#44 AR-1268-4

R.T.: 9.527 min
Delta R.T.: 0.000 min
Response: 16972703
Conc: 266.81 ng/ml



#44 AR-1268-4


R.T.: 8.336 min
Delta R.T.: 0.000 min
Response: 14773829
Conc: 261.66 ng/ml



#45 AR-1268-5

R.T.: 9.947 min
Delta R.T.: 0.000 min
Response: 109112771
Conc: 265.62 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250



#45 AR-1268-5

R.T.: 8.642 min
Delta R.T.: 0.000 min
Response: 95443492
Conc: 250.09 ng/ml