

Surrogate Summary

SDG No.: Q1383

Client: Nobis Group

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PP069264.D	PIBLK-PP069264.D	Tetrachloro-m-xylene	1	20	25.7	128		60	140
		Decachlorobiphenyl	1	20	25.4	127		60	140
		Tetrachloro-m-xylene	2	20	26.0	130		60	140
		Decachlorobiphenyl	2	20	25.4	127		60	140
I.BLK-PP069861.D	PIBLK-PP069861.D	Tetrachloro-m-xylene	1	20	23.1	115		60	140
		Decachlorobiphenyl	1	20	22.2	111		60	140
		Tetrachloro-m-xylene	2	20	22.5	112		60	140
		Decachlorobiphenyl	2	20	22.1	111		60	140
PB166774BL	PB166774BL	Tetrachloro-m-xylene	1	20	23.9	119		44	130
		Decachlorobiphenyl	1	20	23.1	115		60	125
		Tetrachloro-m-xylene	2	20	23.2	116		44	130
		Decachlorobiphenyl	2	20	22.8	114		60	125
PB166774BS	PB166774BS	Tetrachloro-m-xylene	1	20	25.8	129		44	130
		Decachlorobiphenyl	1	20	25.3	127	*	60	125
		Tetrachloro-m-xylene	2	20	24.2	121		44	130
		Decachlorobiphenyl	2	20	23.9	119		60	125
Q1383-01	OU4-PCS-TC-11-021725	Tetrachloro-m-xylene	1	20	25.6	128		44	130
		Decachlorobiphenyl	1	20	21.5	108		60	125
		Tetrachloro-m-xylene	2	20	24.7	124		44	130
		Decachlorobiphenyl	2	20	21.4	107		60	125
Q1383-01MS	OU4-PCS-TC-11-021725MS	Tetrachloro-m-xylene	1	20	25.5	127		44	130
		Decachlorobiphenyl	1	20	20.5	102		60	125
		Tetrachloro-m-xylene	2	20	24.4	122		44	130
		Decachlorobiphenyl	2	20	20.3	101		60	125
Q1383-01MSD	OU4-PCS-TC-11-021725MSD	Tetrachloro-m-xylene	1	20	26.7	134	*	44	130
		Decachlorobiphenyl	1	20	21.4	107		60	125
		Tetrachloro-m-xylene	2	20	24.6	123		44	130
		Decachlorobiphenyl	2	20	21.3	106		60	125
Q1383-09	OU4-PCS-TC-15-021725	Tetrachloro-m-xylene	1	20	26.6	133	*	44	130
		Decachlorobiphenyl	1	20	23.9	119		60	125
		Tetrachloro-m-xylene	2	20	26.6	133	*	44	130
		Decachlorobiphenyl	2	20	23.6	118		60	125
I.BLK-PP069876.D	PIBLK-PP069876.D	Tetrachloro-m-xylene	1	20	23.3	116		60	140
		Decachlorobiphenyl	1	20	22.6	113		60	140
		Tetrachloro-m-xylene	2	20	23.0	115		60	140
		Decachlorobiphenyl	2	20	22.6	113		60	140
Q1383-13	OU4-PCS-TC-17-021725	Tetrachloro-m-xylene	1	20	27.1	136	*	44	130
		Decachlorobiphenyl	1	20	22.4	112		60	125
		Tetrachloro-m-xylene	2	20	26.5	132	*	44	130
		Decachlorobiphenyl	2	20	22.1	111		60	125
Q1383-15	OU4-PCS-TC-18-021725	Tetrachloro-m-xylene	1	20	24.8	124		44	130

Surrogate Summary

SDG No.: Q1383

Client: Nobis Group

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
Q1383-15	OU4-PCS-TC-18-021725	Decachlorobiphenyl	1	20	19.1	96		60	125
		Tetrachloro-m-xylene	2	20	24.8	124		44	130
		Decachlorobiphenyl	2	20	19.7	98		60	125
Q1383-17	OU4-PCS-TC-19-021725	Tetrachloro-m-xylene	1	20	26.5	133	*	44	130
		Decachlorobiphenyl	1	20	22.1	111		60	125
		Tetrachloro-m-xylene	2	20	26.2	131	*	44	130
		Decachlorobiphenyl	2	20	22.0	110		60	125
Q1383-19	OU4-PCS-TC-20-021725	Tetrachloro-m-xylene	1	20	25.7	128		44	130
		Decachlorobiphenyl	1	20	22.2	111		60	125
		Tetrachloro-m-xylene	2	20	26.1	130		44	130
		Decachlorobiphenyl	2	20	21.5	108		60	125
Q1383-21	OU4-CF-15-021725	Tetrachloro-m-xylene	1	20	26.7	134	*	44	130
		Decachlorobiphenyl	1	20	23.2	116		60	125
		Tetrachloro-m-xylene	2	20	26.7	134	*	44	130
		Decachlorobiphenyl	2	20	23.9	119		60	125
I.BLK-PP069889.D	PIBLK-PP069889.D	Tetrachloro-m-xylene	1	20	23.9	120		60	140
		Decachlorobiphenyl	1	20	23.2	116		60	140
		Tetrachloro-m-xylene	2	20	23.5	117		60	140
		Decachlorobiphenyl	2	20	22.4	112		60	140
I.BLK-PP069936.D	PIBLK-PP069936.D	Tetrachloro-m-xylene	1	20	24.4	122		60	140
		Decachlorobiphenyl	1	20	24.0	120		60	140
		Tetrachloro-m-xylene	2	20	24.2	121		60	140
		Decachlorobiphenyl	2	20	23.2	116		60	140
Q1383-09RE	OU4-PCS-TC-15-021725RE	Tetrachloro-m-xylene	1	20	27.8	139	*	44	130
		Decachlorobiphenyl	1	20	25.0	125		60	125
		Tetrachloro-m-xylene	2	20	27.4	137	*	44	130
		Decachlorobiphenyl	2	20	23.9	119		60	125
Q1383-13RE	OU4-PCS-TC-17-021725RE	Tetrachloro-m-xylene	1	20	27.5	138	*	44	130
		Decachlorobiphenyl	1	20	22.9	115		60	125
		Tetrachloro-m-xylene	2	20	26.6	133	*	44	130
		Decachlorobiphenyl	2	20	23.4	117		60	125
Q1383-17RE	OU4-PCS-TC-19-021725RE	Tetrachloro-m-xylene	1	20	27.2	136	*	44	130
		Decachlorobiphenyl	1	20	22.6	113		60	125
		Tetrachloro-m-xylene	2	20	26.5	132	*	44	130
		Decachlorobiphenyl	2	20	23.6	118		60	125
Q1383-21RE	OU4-CF-15-021725RE	Tetrachloro-m-xylene	1	20	27.6	138	*	44	130
		Decachlorobiphenyl	1	20	23.9	119		60	125
		Tetrachloro-m-xylene	2	20	27.1	136	*	44	130
		Decachlorobiphenyl	2	20	23.9	119		60	125
Q1383-11	OU4-PCS-TC-16-021725	Tetrachloro-m-xylene	1	20	25.6	128		44	130
		Decachlorobiphenyl	1	20	23.1	115		60	125

Surrogate Summary

SDG No.: Q1383

Client: Nobis Group

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
Q1383-11	OU4-PCS-TC-16-021725	Tetrachloro-m-xylene	2	20	25.3	126		44	130
		Decachlorobiphenyl	2	20	23.5	117		60	125
I.BLK-PP069951.D	PIBLK-PP069951.D	Tetrachloro-m-xylene	1	20	25.6	128		60	140
		Decachlorobiphenyl	1	20	24.6	123		60	140
		Tetrachloro-m-xylene	2	20	25.5	127		60	140
		Decachlorobiphenyl	2	20	23.8	119		60	140
I.BLK-PP069957.D	PIBLK-PP069957.D	Tetrachloro-m-xylene	1	20	22.5	113		60	140
		Decachlorobiphenyl	1	20	22.3	112		60	140
		Tetrachloro-m-xylene	2	20	22.2	111		60	140
		Decachlorobiphenyl	2	20	22.6	113		60	140
Q1383-03	OU4-PCS-TC-12-021725	Tetrachloro-m-xylene	1	20	23.7	119		44	130
		Decachlorobiphenyl	1	20	21.3	106		60	125
		Tetrachloro-m-xylene	2	20	24.2	121		44	130
		Decachlorobiphenyl	2	20	21.0	105		60	125
Q1383-05	OU4-PCS-TC-13-021725	Tetrachloro-m-xylene	1	20	24.4	122		44	130
		Decachlorobiphenyl	1	20	21.8	109		60	125
		Tetrachloro-m-xylene	2	20	24.2	121		44	130
		Decachlorobiphenyl	2	20	20.4	102		60	125
Q1383-07	OU4-PCS-TC-14-021725	Tetrachloro-m-xylene	1	20	25.6	128		44	130
		Decachlorobiphenyl	1	20	23.2	116		60	125
		Tetrachloro-m-xylene	2	20	25.8	129		44	130
		Decachlorobiphenyl	2	20	22.8	114		60	125
I.BLK-PP069971.D	PIBLK-PP069971.D	Tetrachloro-m-xylene	1	20	21.9	110		60	140
		Decachlorobiphenyl	1	20	21.7	108		60	140
		Tetrachloro-m-xylene	2	20	22.5	112		60	140
		Decachlorobiphenyl	2	20	20.3	101		60	140