

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL012125

Review By	Abdul	Review On	1/22/2025 8:17:54 AM
Supervise By	Ankita	Supervise On	1/22/2025 8:30:39 AM
SubDirectory	PL012125	HP Acquire Method	HP Processing Method pl012125 8081

STD. NAME	STD REF.#
Tune/Reschk	PP23793,PP24095
Initial Calibration Stds	PP23673,PP23674,PP23675,PP23676,PP23677,PP23678,PP23679,PP23680,PP23681,PP23682,PP23683
CCC	PP23686,PP23690,PP23695
Internal Standard/PEM	
ICV/I.BLK	PP23687,PP23693,PP23698
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PL093724.D	21 Jan 2025 10:03		AR\AJ	Ok
2	I.BLK	I.BLK	PL093725.D	21 Jan 2025 10:16		AR\AJ	Ok
3	PEM	PEM	PL093726.D	21 Jan 2025 10:30		AR\AJ	Ok,M
4	RESCHK	RESCHK	PL093727.D	21 Jan 2025 10:43		AR\AJ	Ok
5	PSTDICC100	PSTDICC100	PL093728.D	21 Jan 2025 10:57		AR\AJ	Ok
6	PSTDICC075	PSTDICC075	PL093729.D	21 Jan 2025 11:10		AR\AJ	Ok
7	PSTDICC050	PSTDICC050	PL093730.D	21 Jan 2025 11:24		AR\AJ	Ok
8	PSTDICC025	PSTDICC025	PL093731.D	21 Jan 2025 11:38		AR\AJ	Ok
9	PSTDICC005	PSTDICC005	PL093732.D	21 Jan 2025 11:51		AR\AJ	Ok
10	PCHLORICC1000	PCHLORICC1000	PL093733.D	21 Jan 2025 12:05		AR\AJ	Ok
11	PCHLORICC750	PCHLORICC750	PL093734.D	21 Jan 2025 12:18		AR\AJ	Ok
12	PCHLORICC500	PCHLORICC500	PL093735.D	21 Jan 2025 12:32		AR\AJ	Ok
13	PCHLORICC250	PCHLORICC250	PL093736.D	21 Jan 2025 12:45		AR\AJ	Ok
14	PCHLORICC050	PCHLORICC050	PL093737.D	21 Jan 2025 12:59		AR\AJ	Ok
15	PTOXICC1000	PTOXICC1000	PL093738.D	21 Jan 2025 13:12		AR\AJ	Ok
16	PTOXICC750	PTOXICC750	PL093739.D	21 Jan 2025 13:26		AR\AJ	Ok
17	PTOXICC500	PTOXICC500	PL093740.D	21 Jan 2025 13:39		AR\AJ	Ok
18	PTOXICC250	PTOXICC250	PL093741.D	21 Jan 2025 13:53		AR\AJ	Ok

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL012125

Review By	Abdul	Review On	1/22/2025 8:17:54 AM
Supervise By	Ankita	Supervise On	1/22/2025 8:30:39 AM
SubDirectory	PL012125	HP Acquire Method	HP Processing Method pl012125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP23793,PP24095		
Initial Calibration Stds	PP23673,PP23674,PP23675,PP23676,PP23677,PP23678,PP23679,PP23680,PP23681,PP23682,PP23683		
CCC	PP23686,PP23690,PP23695		
Internal Standard/PEM			
ICV/I.BLK	PP23687,PP23693,PP23698		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PTOXICC100	PTOXICC100	PL093742.D	21 Jan 2025 14:07		AR\AJ	Ok
20	PSTDICV050	ICVPL012125	PL093743.D	21 Jan 2025 14:20		AR\AJ	Ok
21	PCHLORICV500	ICVPL012125CHLOR	PL093744.D	21 Jan 2025 14:47		AR\AJ	Ok,M
22	PTOXICV500	ICVPL012125TOX	PL093745.D	21 Jan 2025 15:14		AR\AJ	Ok
23	I.BLK	I.BLK	PL093746.D	21 Jan 2025 15:41		AR\AJ	Ok
24	PEM	PEM	PL093747.D	21 Jan 2025 15:54		AR\AJ	Ok,M
25	PSTDCCC050	PSTDCCC050	PL093748.D	21 Jan 2025 16:07		AR\AJ	Ok,M
26	Q1093-01RE	RT-3407RE	PL093749.D	21 Jan 2025 16:21	F Flag coming , DCB high in 2nd column	AR\AJ	Confirms
27	I.BLK	I.BLK	PL093750.D	21 Jan 2025 16:34		AR\AJ	Ok
28	PSTDCCC050	PSTDCCC050	PL093751.D	21 Jan 2025 16:48		AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL021925

Review By	Abdul	Review On	2/20/2025 8:45:47 AM
Supervise By	Ankita	Supervise On	2/20/2025 12:59:53 PM
SubDirectory	PL021925	HP Acquire Method	HP Processing Method pl012125 8081

STD. NAME	STD REF.#
Tune/Reschk	PP23793,PP24095
Initial Calibration Stds	PP23673,PP23674,PP23675,PP23676,PP23677,PP23678,PP23679,PP23680,PP23681,PP23682,PP23683
CCC	PP23686,PP23690,PP23695
Internal Standard/PEM	
ICV/I.BLK	PP23687,PP23693,PP23698
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PL094254.D	19 Feb 2025 11:12		AR\AJ	Ok
2	I.BLK	I.BLK	PL094255.D	19 Feb 2025 11:26	DCB high in 2nd column	AR\AJ	Ok
3	PEM	PEM	PL094256.D	19 Feb 2025 11:39		AR\AJ	Ok,M
4	PSTDCCC050	PSTDCCC050	PL094257.D	19 Feb 2025 12:21		AR\AJ	Ok
5	PCHLORCCC500	PCHLORCCC500	PL094258.D	19 Feb 2025 13:39		AR\AJ	Ok,M
6	PTOXCCC500	PTOXCCC500	PL094259.D	19 Feb 2025 14:22		AR\AJ	Ok
7	PB166771BL	PB166771BL	PL094260.D	19 Feb 2025 14:37		AR\AJ	Ok
8	PB166771BS	PB166771BS	PL094261.D	19 Feb 2025 14:51		AR\AJ	Ok,M
9	PB166771BS	PB166771BS	PL094262.D	19 Feb 2025 15:09		AR\AJ	Ok,M
10	PB166771BS	PB166771BS	PL094263.D	19 Feb 2025 15:27		AR\AJ	Ok
11	Q1382-01	OU4-PCS-TC-01-02172	PL094264.D	19 Feb 2025 16:12		AR\AJ	Ok,M
12	Q1382-01MS	OU4-PCS-TC-01-02172	PL094265.D	19 Feb 2025 16:25		AR\AJ	Ok,M
13	Q1382-01MSD	OU4-PCS-TC-01-02172	PL094266.D	19 Feb 2025 16:39		AR\AJ	Ok,M
14	Q1382-03	OU4-PCS-TC-02-02172	PL094267.D	19 Feb 2025 16:52		AR\AJ	Ok,M
15	Q1382-05	OU4-PCS-TC-03-02172	PL094268.D	19 Feb 2025 17:06		AR\AJ	Ok,M
16	Q1382-07	OU4-PCS-TC-04-02172	PL094269.D	19 Feb 2025 17:19		AR\AJ	Ok,M
17	I.BLK	I.BLK	PL094270.D	19 Feb 2025 17:33		AR\AJ	Ok
18	PSTDCCC050	PSTDCCC050	PL094271.D	19 Feb 2025 17:47		AR\AJ	Ok,M

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL021925

Review By	Abdul	Review On	2/20/2025 8:45:47 AM
Supervise By	Ankita	Supervise On	2/20/2025 12:59:53 PM
SubDirectory	PL021925	HP Acquire Method	HP Processing Method pl012125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP23793,PP24095		
Initial Calibration Stds	PP23673,PP23674,PP23675,PP23676,PP23677,PP23678,PP23679,PP23680,PP23681,PP23682,PP23683		
CCC	PP23686,PP23690,PP23695		
Internal Standard/PEM			
ICV/I.BLK	PP23687,PP23693,PP23698		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PCHLORCCC500	PCHLORCCC500	PL094272.D	19 Feb 2025 18:00		AR\AJ	Ok,M
20	PTOXCCC500	PTOXCCC500	PL094273.D	19 Feb 2025 18:14		AR\AJ	Ok,M
21	Q1382-09	OU4-PCS-TC-05-02172	PL094274.D	19 Feb 2025 18:27	DCB high in 2nd column	AR\AJ	Ok,M
22	Q1382-11	OU4-PCS-TC-06-02172	PL094275.D	19 Feb 2025 18:41		AR\AJ	Ok,M
23	Q1382-13	OU4-PCS-TC-07-02172	PL094276.D	19 Feb 2025 18:54		AR\AJ	Ok,M
24	Q1382-15	OU4-PCS-TC-08-02172	PL094277.D	19 Feb 2025 19:08		AR\AJ	Ok,M
25	Q1382-17	OU4-PCS-TC-09-02172	PL094278.D	19 Feb 2025 19:21		AR\AJ	Ok,M
26	Q1382-19	OU4-PCS-TC-10-02172	PL094279.D	19 Feb 2025 19:35		AR\AJ	Ok,M
27	Q1384-01	OR-03-021823025	PL094280.D	19 Feb 2025 19:49		AR\AJ	Ok,M
28	Q1385-01	ST-1	PL094281.D	19 Feb 2025 20:02		AR\AJ	Ok
29	Q1383-01	OU4-PCS-TC-11-02172	PL094282.D	19 Feb 2025 20:16		AR\AJ	Ok,M
30	Q1383-01MS	OU4-PCS-TC-11-02172	PL094283.D	19 Feb 2025 20:29		AR\AJ	Ok,M
31	I.BLK	I.BLK	PL094284.D	19 Feb 2025 20:43		AR\AJ	Ok
32	PEM	PEM	PL094285.D	19 Feb 2025 20:57		AR\AJ	Ok,M
33	PSTDCCC050	PSTDCCC050	PL094286.D	19 Feb 2025 21:10		AR\AJ	Ok,M
34	Q1383-01MSD	OU4-PCS-TC-11-02172	PL094287.D	20 Feb 2025 08:29		AR\AJ	Ok,M
35	I.BLK	I.BLK	PL094288.D	20 Feb 2025 08:43		AR\AJ	Ok
36	PSTDCCC050	PSTDCCC050	PL094289.D	20 Feb 2025 08:56		AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL022025

Review By	Abdul	Review On	2/21/2025 7:04:00 AM
Supervise By	Ankita	Supervise On	2/21/2025 11:11:13 AM
SubDirectory	PL022025	HP Acquire Method	HP Processing Method pl012125 8081

STD. NAME	STD REF.#
Tune/Reschk	PP23793,PP24095
Initial Calibration Stds	PP23673,PP23674,PP23675,PP23676,PP23677,PP23678,PP23679,PP23680,PP23681,PP23682,PP23683
CCC	PP23686,PP23690,PP23695
Internal Standard/PEM	
ICV/I.BLK	PP23687,PP23693,PP23698
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PL094290.D	20 Feb 2025 09:28		AR\AJ	Ok
2	I.BLK	I.BLK	PL094291.D	20 Feb 2025 09:42		AR\AJ	Ok
3	PEM	PEM	PL094292.D	20 Feb 2025 10:02		AR\AJ	Ok,M
4	PSTDCCC050	PSTDCCC050	PL094293.D	20 Feb 2025 10:16		AR\AJ	Ok,M
5	Q1383-03	OU4-PCS-TC-12-02172	PL094294.D	20 Feb 2025 10:29	TCMX high in 2nd column	AR\AJ	Ok,M
6	Q1383-05	OU4-PCS-TC-13-02172	PL094295.D	20 Feb 2025 10:43	TCMX high in 2nd column	AR\AJ	Ok,M
7	Q1383-07	OU4-PCS-TC-14-02172	PL094296.D	20 Feb 2025 10:57	TCMX high in 2nd column	AR\AJ	Ok,M
8	Q1383-09	OU4-PCS-TC-15-02172	PL094297.D	20 Feb 2025 11:11		AR\AJ	Ok,M
9	Q1383-11	OU4-PCS-TC-16-02172	PL094298.D	20 Feb 2025 11:24	TCMX high in 2nd column	AR\AJ	Ok,M
10	Q1383-13	OU4-PCS-TC-17-02172	PL094299.D	20 Feb 2025 11:38		AR\AJ	Ok,M
11	Q1383-15	OU4-PCS-TC-18-02172	PL094300.D	20 Feb 2025 11:52		AR\AJ	Ok,M
12	Q1383-17	OU4-PCS-TC-19-02172	PL094301.D	20 Feb 2025 12:06		AR\AJ	Ok,M
13	Q1383-19	OU4-PCS-TC-20-02172	PL094302.D	20 Feb 2025 12:19		AR\AJ	Ok,M
14	Q1383-21	OU4-CF-15-021725	PL094303.D	20 Feb 2025 12:33		AR\AJ	Ok,M
15	I.BLK	I.BLK	PL094304.D	20 Feb 2025 12:47		AR\AJ	Ok
16	PSTDCCC050	PSTDCCC050	PL094305.D	20 Feb 2025 13:00		AR\AJ	Ok,M
17	PCHLORCCC500	PCHLORCCC500	PL094306.D	20 Feb 2025 14:39		AR\AJ	Ok,M
18	PTOXCCC500	PTOXCCC500	PL094307.D	20 Feb 2025 15:53		AR\AJ	Ok,M

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL022025

Review By	Abdul	Review On	2/21/2025 7:04:00 AM
Supervise By	Ankita	Supervise On	2/21/2025 11:11:13 AM
SubDirectory	PL022025	HP Acquire Method	HP Processing Method pl012125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP23793,PP24095		
Initial Calibration Stds	PP23673,PP23674,PP23675,PP23676,PP23677,PP23678,PP23679,PP23680,PP23681,PP23682,PP23683		
CCC	PP23686,PP23690,PP23695		
Internal Standard/PEM			
ICV/I.BLK	PP23687,PP23693,PP23698		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PB166775BL	PB166775BL	PL094308.D	20 Feb 2025 16:08		AR\AJ	Ok
20	PB166775BS	PB166775BS	PL094309.D	20 Feb 2025 16:21		AR\AJ	Ok,M
21	PB166775BS	PB166775BS	PL094310.D	20 Feb 2025 16:35		AR\AJ	Ok,M
22	PB166775BS	PB166775BS	PL094311.D	20 Feb 2025 17:43		AR\AJ	Ok,M
23	I.BLK	I.BLK	PL094312.D	20 Feb 2025 17:58		AR\AJ	Ok
24	PSTDCCC050	PSTDCCC050	PL094313.D	20 Feb 2025 18:12	CCC failing high for com#16,20	AR\AJ	Ok,M
25	PCHLORCCC500	PCHLORCCC500	PL094314.D	20 Feb 2025 18:26		AR\AJ	Ok,M
26	PTOXCCC500	PTOXCCC500	PL094315.D	20 Feb 2025 18:39		AR\AJ	Ok,M
27	PB166800BL	PB166800BL	PL094316.D	20 Feb 2025 18:53		AR\AJ	Ok
28	PB166800BS	PB166800BS	PL094317.D	20 Feb 2025 19:06		AR\AJ	Ok,M
29	PB166800BSD	PB166800BSD	PL094318.D	20 Feb 2025 19:20		AR\AJ	Ok,M
30	Q1393-05	PURGE-WATER-COMF	PL094319.D	20 Feb 2025 19:34	F Flag coming , TCMX low in 1st column	AR\AJ	Not Ok
31	I.BLK	I.BLK	PL094320.D	20 Feb 2025 19:47		AR\AJ	Ok
32	PSTDCCC050	PSTDCCC050	PL094321.D	20 Feb 2025 20:14	Comp#16 Recovery high	AR\AJ	Ok,M

M : Manual Integration