

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY020325

Review By	Mahesh Dadoda	Review On	2/4/2025 3:47:24 PM
Supervise By	Semsettin Yesilyurt	Supervise On	2/4/2025 3:53:22 PM
SubDirectory	VY020325	HP Acquire Method	MSVOA_Y
		HP Processing Method	82y020325s.m

STD. NAME	STD REF.#
Tune/Reschk	VP132839
Initial Calibration Stds	VP132837,VP132841,VP132843,VP132845,VP132847,VP132849
CCC	VP132853
Internal Standard/PEM	VP131783
ICV/I.BLK	VP132851
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY021019.D	03 Feb 2025 09:51		SY/MD	Ok
2	VSTDICC005	VSTDICC005	VY021020.D	03 Feb 2025 10:35	Good for DOD	SY/MD	Ok,M
3	VSTDICC010	VSTDICC010	VY021021.D	03 Feb 2025 10:57		SY/MD	Ok,M
4	VSTDICC020	VSTDICC020	VY021022.D	03 Feb 2025 11:20		SY/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VY021023.D	03 Feb 2025 11:43		SY/MD	Ok,M
6	VSTDICC100	VSTDICC100	VY021024.D	03 Feb 2025 12:21		SY/MD	Ok,M
7	VSTDICC150	VSTDICC150	VY021025.D	03 Feb 2025 12:44		SY/MD	Ok,M
8	VIBLK	VIBLK	VY021026.D	03 Feb 2025 13:11		SY/MD	Ok
9	VSTDICV050	ICVVY020325	VY021027.D	03 Feb 2025 13:37		SY/MD	Ok,M
10	VY0203SBL01	VY0203SBL01	VY021028.D	03 Feb 2025 14:13		SY/MD	Ok
11	VY0203SBS01	VY0203SBS01	VY021029.D	03 Feb 2025 14:55		SY/MD	Ok,M
12	VY0203SBSD01	VY0203SBSD01	VY021030.D	03 Feb 2025 15:17		SY/MD	Ok,M
13	Q1239-10	357	VY021031.D	03 Feb 2025 15:50	vial B Not purged	SY/MD	Not Ok
14	Q1215-01	JPP-29.1-012825	VY021032.D	03 Feb 2025 16:14	vial A ISTD Fail	SY/MD	Ok
15	Q1215-05	JPP-29.2-012825	VY021033.D	03 Feb 2025 16:37	Not purged vial-A	SY/MD	Not Ok
16	Q1216-01	JPP-18.1-012825	VY021034.D	03 Feb 2025 17:01	vial A Internal Standard Fail	SY/MD	ReRun
17	Q1216-05	JPP-21.1-012825	VY021035.D	03 Feb 2025 17:24	vial B ISTD Fail	SY/MD	Ok
18	IBLK	IBLK	VY021036.D	03 Feb 2025 17:47	Not purged	SY/MD	Not Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY020325

Review By	Mahesh Dadoda	Review On	2/4/2025 3:47:24 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	2/4/2025 3:53:22 PM		
SubDirectory	VY020325	HP Acquire Method	MSVOA_Y	HP Processing Method	82y020325s.m
STD. NAME		STD REF.#			
Tune/Reschk		VP132839			
Initial Calibration Stds		VP132837,VP132841,VP132843,VP132845,VP132847,VP132849			
CCC		VP132853			
Internal Standard/PEM		VP131783			
ICV/I.BLK		VP132851			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	IBLK	IBLK	VY021037.D	03 Feb 2025 18:11	Not purged	SY/MD	Not Ok
20	Q1216-17	JPP-26.2-012825	VY021038.D	03 Feb 2025 18:34	vial B Internal Standard Fail	SY/MD	Ok
21	Q1206-01	JPP-20.1-012725	VY021039.D	03 Feb 2025 18:58	vial B	SY/MD	Ok
22	Q1206-05	JPP-16.3-012725	VY021040.D	03 Feb 2025 19:21	vial B	SY/MD	Ok
23	Q1207-01	JPP-2.1-012725	VY021041.D	03 Feb 2025 19:44	vial-B	SY/MD	Ok
24	Q1207-05	JPP-5.1-012725	VY021042.D	03 Feb 2025 20:08	vial-B	SY/MD	Ok
25	Q1207-09	JPP-4.5-012725	VY021043.D	03 Feb 2025 20:31	vial-B	SY/MD	Ok
26	VSTDCCC050	VSTDCCC050EC	VY021044.D	03 Feb 2025 20:54		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY022125

Review By	Maresh Dadoda	Review On	2/24/2025 1:32:08 PM
Supervise By	Semsettin Yesilyurt	Supervise On	2/24/2025 1:34:10 PM
SubDirectory	VY022125	HP Acquire Method	MSVOA_Y HP Processing Method 82y020325s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133100
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133101,VP133102 VP131783

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY021274.D	21 Feb 2025 09:06		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY021275.D	21 Feb 2025 09:37		SY/MD	Ok,M
3	VY0221SBL01	VY0221SBL01	VY021276.D	21 Feb 2025 11:12		SY/MD	Ok
4	VY0221SBS01	VY0221SBS01	VY021277.D	21 Feb 2025 11:43		SY/MD	Ok,M
5	VY0221SBSD01	VY0221SBSD01	VY021278.D	21 Feb 2025 12:06		SY/MD	Ok,M
6	Q1383-01	OU4-PCS-TC-11-02172	VY021279.D	21 Feb 2025 12:41	vial-B	SY/MD	Ok
7	Q1383-03	OU4-PCS-TC-12-02172	VY021280.D	21 Feb 2025 13:04	vial-B	SY/MD	Ok
8	Q1383-05	OU4-PCS-TC-13-02172	VY021281.D	21 Feb 2025 13:27	vial-B	SY/MD	Ok
9	Q1383-07	OU4-PCS-TC-14-02172	VY021282.D	21 Feb 2025 13:51	vial-B	SY/MD	Ok
10	VY0221SBSD02	VY0221SBSD02	VY021283.D	21 Feb 2025 14:14	Not Required	SY/MD	Not Ok
11	Q1383-11	OU4-PCS-TC-16-02172	VY021284.D	21 Feb 2025 14:37	vial-B	SY/MD	Ok
12	Q1383-13	OU4-PCS-TC-17-02172	VY021285.D	21 Feb 2025 15:00	vial-B	SY/MD	Ok
13	Q1383-15	OU4-PCS-TC-18-02172	VY021286.D	21 Feb 2025 15:24	vial-B	SY/MD	Ok
14	Q1383-17	OU4-PCS-TC-19-02172	VY021287.D	21 Feb 2025 15:47	vial-B	SY/MD	Ok
15	Q1383-19	OU4-PCS-TC-20-02172	VY021288.D	21 Feb 2025 16:10	vial-B	SY/MD	Ok
16	Q1383-21	OU4-CF-15-021725	VY021289.D	21 Feb 2025 16:34	vial-B	SY/MD	Ok
17	Q1382-07	OU4-PCS-TC-04-02172	VY021290.D	21 Feb 2025 16:57	vial-B	SY/MD	Ok
18	Q1382-09	OU4-PCS-TC-05-02172	VY021291.D	21 Feb 2025 17:21	vial-B	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY022125

Review By	Mahesh Dadoda	Review On	2/24/2025 1:32:08 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	2/24/2025 1:34:10 PM		
SubDirectory	VY022125	HP Acquire Method	MSVOA_Y	HP Processing Method	82y020325s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP133100			
CCC		VP133101,VP133102			
Internal Standard/PEM		VP131783			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q1382-11	OU4-PCS-TC-06-02172	VY021292.D	21 Feb 2025 17:44	vial-B	SY/MD	Ok
20	Q1382-13	OU4-PCS-TC-07-02172	VY021293.D	21 Feb 2025 18:08	vial-B	SY/MD	Ok
21	Q1382-15	OU4-PCS-TC-08-02172	VY021294.D	21 Feb 2025 18:31	vial-B	SY/MD	Ok
22	Q1382-17	OU4-PCS-TC-09-02172	VY021295.D	21 Feb 2025 18:54	vial-B	SY/MD	Ok
23	Q1388-01RE	72-12016RE	VY021296.D	21 Feb 2025 19:18	vial-B Internal Standard Fail	SY/MD	Confirms
24	Q1393-01	SOIL-COMP	VY021297.D	21 Feb 2025 19:41	vial-B Not purge	SY/MD	Not Ok
25	VSTDCCC050	VSTDCCC050EC	VY021298.D	21 Feb 2025 20:27		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY022525

Review By	Mahesh Dadoda	Review On	2/26/2025 1:45:35 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	2/26/2025 1:46:55 PM		
SubDirectory	VY022525	HP Acquire Method	MSVOA_Y	HP Processing Method	82y022525s.m

STD. NAME	STD REF.#
Tune/Reschk	VP133154
Initial Calibration Stds	VP133137,VP133140,VP133143,VP133146,VP133147,VP133149
CCC	
Internal Standard/PEM	VP131783
ICV/I.BLK	VP133150
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampled	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY021308.D	25 Feb 2025 12:10		SY/MD	Ok
2	VSTDIC005	VSTDIC005	VY021309.D	25 Feb 2025 12:40		SY/MD	Ok,M
3	VSTDIC010	VSTDIC010	VY021310.D	25 Feb 2025 13:03		SY/MD	Ok,M
4	VSTDIC020	VSTDIC020	VY021311.D	25 Feb 2025 13:26	Acrolein fail for method	SY/MD	Ok,M
5	VSTDIC050	VSTDIC050	VY021312.D	25 Feb 2025 14:48		SY/MD	Ok,M
6	VSTDIC100	VSTDIC100	VY021313.D	25 Feb 2025 15:34	Not used	SY/MD	Not Ok
7	VSTDIC150	VSTDIC150	VY021314.D	25 Feb 2025 15:57		SY/MD	Ok,M
8	VIBLK	VIBLK	VY021315.D	25 Feb 2025 16:24		SY/MD	Ok
9	VSTDIC100	VSTDIC100	VY021316.D	25 Feb 2025 16:49		SY/MD	Ok,M
10	VSTDICV050	ICVVY022525	VY021317.D	25 Feb 2025 17:31	ICV Failed for comp. #10	SY/MD	Ok,M
11	VIBLK	VIBLK	VY021318.D	25 Feb 2025 17:54		SY/MD	Ok

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY022625

Review By	Maresh Dadoda	Review On	2/27/2025 12:12:40 PM
Supervise By	Semsettin Yesilyurt	Supervise On	2/27/2025 12:13:53 PM
SubDirectory	VY022625	HP Acquire Method	MSVOA_Y HP Processing Method 82Y022525S.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133161
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133162,VP133163 VP131783

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY021319.D	26 Feb 2025 09:12		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY021320.D	26 Feb 2025 10:11		SY/MD	Ok,M
3	VY0226SBL01	VY0226SBL01	VY021321.D	26 Feb 2025 10:50		SY/MD	Ok
4	VY0226SBS01	VY0226SBS01	VY021322.D	26 Feb 2025 11:22		SY/MD	Ok,M
5	VY0226SBSD01	VY0226SBSD01	VY021323.D	26 Feb 2025 11:44		SY/MD	Ok,M
6	Q1416-01	OK-01-022425	VY021324.D	26 Feb 2025 12:20	vial-A ISTD Fail	SY/MD	ReRun
7	Q1413-02	BU-2-022125	VY021325.D	26 Feb 2025 12:43	vial-A ISTD Fail	SY/MD	ReRun
8	Q1418-01	TR-06-02242025	VY021326.D	26 Feb 2025 13:07	vial-A ISTD Fail	SY/MD	Ok
9	Q1413-02RE	BU-2-022125RE	VY021327.D	26 Feb 2025 13:30	vial-B ISTD Fail	SY/MD	Confirms
10	Q1401-07	BP-VPB-192-GW-825-8	VY021328.D	26 Feb 2025 13:54	vial-A	SY/MD	Ok
11	Q1401-08	BP-VPB-192-GW-860-8	VY021329.D	26 Feb 2025 14:17	vial-A	SY/MD	Ok
12	Q1401-09	BP-VPB-192-GW-880-8	VY021330.D	26 Feb 2025 14:40	vial-A	SY/MD	Ok
13	Q1382-19	OU4-PCS-TC-10-02172	VY021331.D	26 Feb 2025 15:04	vial-A	SY/MD	Ok
14	Q1383-09	OU4-PCS-TC-15-02172	VY021332.D	26 Feb 2025 15:27	vial-B	SY/MD	Ok
15	Q1411-03	HR-03-02212025	VY021333.D	26 Feb 2025 15:51	vial-A NOT PURGE	SY/MD	Not Ok
16	Q1411-01	HR-02-02212025	VY021334.D	26 Feb 2025 16:14	vial-A NOT PURGE	SY/MD	Not Ok
17	Q1415-02	B-172-SB01	VY021335.D	26 Feb 2025 16:37	vial-A	SY/MD	Ok
18	Q1415-01	B-163-SB01	VY021336.D	26 Feb 2025 17:01	vial-A	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY022625

Review By	Mahesh Dadoda	Review On	2/27/2025 12:12:40 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	2/27/2025 12:13:53 PM		
SubDirectory	VY022625	HP Acquire Method	MSVOA_Y	HP Processing Method	82Y022525S.m
STD. NAME	STD REF.#				
Tune/Reschk Initial Calibration Stds	VP133161				
CCC	VP133162,VP133163				
Internal Standard/PEM	VP131783				
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q1415-04	B-172-SB02	VY021337.D	26 Feb 2025 17:24	vial-A	SY/MD	Ok,M
20	Q1415-03	B-163-SB02	VY021338.D	26 Feb 2025 17:48	vial-A Internal Standard Fail	SY/MD	ReRun
21	Q1420-07	TP-2-WC-VOC	VY021339.D	26 Feb 2025 18:11	vial-A Internal Standard Fail	SY/MD	ReRun
22	Q1420-03	TP-1-WC-VOC	VY021340.D	26 Feb 2025 18:34	vial-A	SY/MD	Ok
23	Q1427-01	VNJ-227	VY021341.D	26 Feb 2025 18:58	vial-A Surrogate Fail	SY/MD	ReRun
24	Q1380-07MS	BP-VPB-192-GW-725-7	VY021342.D	26 Feb 2025 19:21	vial-B	SY/MD	ReRun
25	Q1380-08MSD	BP-VPB-192-GW-725-7	VY021343.D	26 Feb 2025 19:43	vial-B Internal Standard Fail	SY/MD	ReRun
26	VSTDCCC050	VSTDCCC050EC	VY021344.D	26 Feb 2025 20:06		SY/MD	Ok,M

M : Manual Integration