

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: NOBI03
 Lab Code: CHEM Case No.: Q1383 SAS No.: Q1383 SDG NO.: Q1383
 Lab File ID: VY021320.D Date Analyzed: 02/26/2025
 Instrument ID: MSVOA_Y Time Analyzed: 10:11
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	193888	7.71	308714	8.62	268805	11.42
UPPER LIMIT	387776	8.213	617428	9.116	537610	11.92
LOWER LIMIT	96944	7.213	154357	8.116	134403	10.92
EPA SAMPLE NO.						
OU4-PCS-TC-15-021725	240886	7.71	477609	8.62	478629	11.41
VY0226SBL01	251129	7.71	483695	8.62	470141	11.42
VY0226SBS01	184417	7.71	291138	8.62	254891	11.41
VY0226SBSD01	186200	7.71	292071	8.62	258744	11.41

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.