

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: NOBI03  
 Lab Code: CHEM Case No.: Q1383 SAS No.: Q1383 SDG NO.: Q1383  
 Lab File ID: VY021275.D Date Analyzed: 02/21/2025  
 Instrument ID: MSVOA\_Y Time Analyzed: 09:37  
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

|                      | IS4<br>AREA # | RT #   |  |  |  |  |
|----------------------|---------------|--------|--|--|--|--|
| 12 HOUR STD          | 91135         | 13.359 |  |  |  |  |
| UPPER LIMIT          | 182270        | 13.859 |  |  |  |  |
| LOWER LIMIT          | 45567.5       | 12.859 |  |  |  |  |
| EPA SAMPLE NO.       |               |        |  |  |  |  |
| OU4-PCS-TC-11-021725 | 110482        | 13.36  |  |  |  |  |
| OU4-PCS-TC-12-021725 | 115459        | 13.35  |  |  |  |  |
| OU4-PCS-TC-13-021725 | 125048        | 13.35  |  |  |  |  |
| OU4-PCS-TC-14-021725 | 125321        | 13.35  |  |  |  |  |
| OU4-PCS-TC-16-021725 | 118737        | 13.35  |  |  |  |  |
| OU4-PCS-TC-17-021725 | 121669        | 13.35  |  |  |  |  |
| OU4-PCS-TC-18-021725 | 126536        | 13.35  |  |  |  |  |
| OU4-PCS-TC-19-021725 | 120303        | 13.35  |  |  |  |  |
| OU4-PCS-TC-20-021725 | 130553        | 13.35  |  |  |  |  |
| OU4-CF-15-021725     | 125058        | 13.35  |  |  |  |  |
| VY0221SBL01          | 107899        | 13.35  |  |  |  |  |
| VY0221SBS01          | 99790         | 13.36  |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.