

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: NOBI03
Lab Code: CHEM Case No.: Q1383 SAS No.: Q1383 SDG NO.: Q1383
Lab File ID: VY021319.D BFB Injection Date: 02/26/2025
Instrument ID: MSVOA_Y BFB Injection Time: 09:12
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	21.5
75	30.0 - 60.0% of mass 95	55.4
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.6
173	Less than 2.0% of mass 174	1.2 (1.5) 1
174	50.0 - 100.0% of mass 95	78.5
175	5.0 - 9.0% of mass 174	6.5 (8.3) 1
176	95.0 - 101.0% of mass 174	75.9 (96.7) 1
177	5.0 - 9.0% of mass 176	5.1 (6.7) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VY021320.D	02/26/2025	10:11
VY0226SBL01	VY0226SBL01	VY021321.D	02/26/2025	10:50
VY0226SBS01	VY0226SBS01	VY021322.D	02/26/2025	11:22
VY0226SBSD01	VY0226SBSD01	VY021323.D	02/26/2025	11:44
OU4-PCS-TC-15-021725	Q1383-09	VY021332.D	02/26/2025	15:27
VSTDCCC050EC	VSTDCCC050	VY021344.D	02/26/2025	20:06