

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: GFEL01

Lab Code: CHEM Case No.: Q1386 SAS No.: Q1386 SDG No.: Q1386

Instrument ID: MSVOA_L Calibration Date/Time: 02/21/2025 10:23

Lab File ID: VL042034.D Init. Calib. Date(s): 02/14/2025 02/14/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 12:21 15:31

GC Column: RTX-1 ID: 0.32 (mm)

COMPOUND	RRF	RRF010	MIN RRF	%D	MAX%D
Vinyl Chloride	0.800	0.586		-26.75	30
1,1-Dichloroethene	0.646	0.537		-16.87	30
cis-1,2-Dichloroethene	1.797	2.017		12.24	30
1,1,1-Trichloroethane	2.920	2.904		-0.55	30
Trichloroethene	0.492	0.559		13.62	30
Tetrachloroethene	0.423	0.448		5.91	30
1-Bromo-4-Fluorobenzene	0.825	0.832		0.85	30

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.