

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GFEL01
 Lab Code: CHEM Case No.: Q1386 SAS No.: Q1386 SDG NO.: Q1386
 Lab File ID: VL042034.D Date Analyzed: 02/21/2025
 Instrument ID: MSVOA_L Time Analyzed: 10:23
 GC Column: RTX-1 ID: 0.32 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	96177	2.78	241609	3.96	222442	8.88
UPPER LIMIT	134648	3.11	338253	4.29	311419	9.21
LOWER LIMIT	57706.2	2.45	144965	3.63	133465	8.55
EPA SAMPLE NO.						
IA1A	89356	2.78	243432	3.96	217289	8.88
IA1ADUP	102359	2.79	260637	3.96	236485	8.89
VL0221ABL01	104350	2.79	269157	3.96	235545	8.88
VL0221ABS01	87230	2.78	234490	3.95	217352	8.88

IS1 = Bromochloromethane
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +40% of internal standard area
 AREA LOWER LIMIT = -40% of internal standard area
 RT UPPER LIMIT = +0.33 minutes of internal standard RT
 RT LOWER LIMIT = -0.33 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.