

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q1387

Project: Monthly 2025

Run No.: LB134761

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134761BS							
BOD5	mg/L	198	224		113	1	84.6-115.4	02/20/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q1387

Project: Monthly 2025

Run No.: LB134768

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134768BS							
TPH	mg/L	20.0	17.0		85	1	78-114	02/21/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q1387

Project: Monthly 2025

Run No.: LB134768

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134768BSD							
TPH	mg/L	20.0	17.3		86	1	78-114	02/21/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q1387

Project: Monthly 2025

Run No.: LB134784

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134784BS							
TSS	mg/L	550	532		97	1	90-110	02/21/2025