

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN021025

Review By	Rahul	Review On	2/11/2025 4:21:36 PM
Supervise By	Jagrut	Supervise On	2/11/2025 6:22:45 PM
SubDirectory	BN021025	HP Acquire Method	BNA_N, 8270_HP Processing Method bn021025

STD. NAME	STD REF.#
Tune/Reschk	SP6717
Initial Calibration Stds	SP6738,SP6736,SP6735,SP6734,SP6733,SP6732,SP6731
CCC	SP6735
Internal Standard/PEM	SP6682,1ul/100ul sample
ICV/I.BLK	SP6684
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN036408.D	10 Feb 2025 11:46		RC/JU	Ok
2	SSTDICC0.1	SSTDICC0.1	BN036409.D	10 Feb 2025 12:25		RC/JU	Ok
3	SSTDICC0.2	SSTDICC0.2	BN036410.D	10 Feb 2025 13:01		RC/JU	Ok
4	SSTDICCC0.4	SSTDICCC0.4	BN036411.D	10 Feb 2025 13:36	The Calibration is Good For DOD	RC/JU	Ok
5	SSTDICC0.8	SSTDICC0.8	BN036412.D	10 Feb 2025 14:12		RC/JU	Ok
6	SSTDICC1.6	SSTDICC1.6	BN036413.D	10 Feb 2025 14:48		RC/JU	Ok
7	SSTDICC3.2	SSTDICC3.2	BN036414.D	10 Feb 2025 15:24		RC/JU	Ok
8	SSTDICC5.0	SSTDICC5.0	BN036415.D	10 Feb 2025 16:00	Compound #20 removed from 5.0ppm	RC/JU	Ok
9	SSTDICV0.4	ICVBN021025	BN036416.D	10 Feb 2025 16:36		RC/JU	Ok
10	PB166297BL	PB166297BL	BN036417.D	10 Feb 2025 17:48	Analyzed for contamination check	RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN022425

Review By	Rahul	Review On	2/25/2025 8:55:19 AM
Supervise By	mohammad	Supervise On	2/28/2025 7:32:16 AM
SubDirectory	BN022425	HP Acquire Method	BNA_N, 8270_HP Processing Method bn021025

STD. NAME	STD REF.#
Tune/Reschk	SP6717
Initial Calibration Stds	SP6738,SP6736,SP6735,SP6734,SP6733,SP6732,SP6731
CCC	SP6735
Internal Standard/PEM	SP6740,1ul/100ul sample
ICV/I.BLK	SP6684
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN036491.D	24 Feb 2025 10:10		RC/JU	Ok
2	SSTDCCC0.4	SSTDCCC0.4	BN036492.D	24 Feb 2025 10:49		RC/JU	Ok
3	PB166832BL	PB166832BL	BN036493.D	24 Feb 2025 11:25		RC/JU	Ok
4	Q1401-02	BP-VPB-192-GW-825-8	BN036494.D	24 Feb 2025 12:01		RC/JU	Ok
5	Q1401-04	BP-VPB-192-GW-860-8	BN036495.D	24 Feb 2025 12:37	Internal Standard Fail	RC/JU	ReRun
6	Q1401-06	BP-VPB-192-GW-900-9	BN036496.D	24 Feb 2025 13:13		RC/JU	Ok,M
7	Q1402-01	RW5-SP100-20250219	BN036497.D	24 Feb 2025 13:49	Need 5X Dilution - Not used.	RC/JU	Not Ok
8	Q1402-02	RW5-SP201-20250219	BN036498.D	24 Feb 2025 14:25		RC/JU	Ok
9	Q1402-03	RW5-SP303-20250219	BN036499.D	24 Feb 2025 15:01		RC/JU	Ok
10	Q1401-04RE	BP-VPB-192-GW-860-8	BN036500.D	24 Feb 2025 15:37	Internal Standard Fail	RC/JU	Confirms
11	Q1402-01DL	RW5-SP100-20250219	BN036501.D	24 Feb 2025 16:13	Dilution not matching With First Run	RC/JU	Not Ok
12	PB166832BS	PB166832BS	BN036502.D	24 Feb 2025 16:49		RC/JU	Ok,M
13	Q1402-01	RW5-SP100-20250219	BN036503.D	24 Feb 2025 18:14	Need 5X Dilution	RC/JU	Dilution
14	Q1402-01DL	RW5-SP100-20250219	BN036504.D	24 Feb 2025 18:50		RC/JU	Ok
15	PB166832BSD	PB166832BSD	BN036505.D	24 Feb 2025 19:25		RC/JU	Ok,M
16	SSTDCCC0.4	SSTDCCC0.4EC	BN036506.D	24 Feb 2025 20:02		RC/JU	Ok,M

M : Manual Integration