

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM

SAS No.: Q1401

SDG NO.: Q1401

Lab File ID: BN036491.D

DFTPP Injection Date: 02/24/2025

Instrument ID: BNA_N

DFTPP Injection Time: 10:10

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	52.3
68	Less than 2.0% of mass 69	0.5 (1.1) 1
69	Mass 69 relative abundance	47.7
70	Less than 2.0% of mass 69	0.3 (0.6) 1
127	10.0 - 80.0% of mass 198	47.5
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7.1
275	10.0 - 60.0% of mass 198	25.9
365	Greater than 1% of mass 198	3.9
441	Present, but less than mass 443	9.5
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	10.8 (18.1) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN036492.D	02/24/2025	10:49
PB166832BL	PB166832BL	BN036493.D	02/24/2025	11:25
BP-VPB-192-GW-825-827	Q1401-02	BN036494.D	02/24/2025	12:01
BP-VPB-192-GW-860-862	Q1401-04	BN036495.D	02/24/2025	12:37
BP-VPB-192-GW-900-902	Q1401-06	BN036496.D	02/24/2025	13:13
BP-VPB-192-GW-860-862RE	Q1401-04RE	BN036500.D	02/24/2025	15:37
PB166832BS	PB166832BS	BN036502.D	02/24/2025	16:49
PB166832BSD	PB166832BSD	BN036505.D	02/24/2025	19:25
SSTDCCC0.4EC	SSTDCCC0.4	BN036506.D	02/24/2025	20:02