

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1401 SAS No.: Q1401 SDG No.: Q1401

Instrument ID: MSVOA_X Calibration Date/Time: 02/25/2025 16:28

Lab File ID: VX045042.D Init. Calib. Date(s): 02/10/2025 02/10/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:25 12:28

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Chloromethane	0.854	0.726	0.1	-14.99	50
Vinyl Chloride	0.827	0.730		-11.73	50
Bromomethane	0.247	0.257		4.05	50
Chloroethane	0.307	0.394		28.34	50
Trichlorofluoromethane	1.046	1.033		-1.24	50
1,1,2-Trichlorotrifluoroethane	0.632	0.618		-2.21	50
1,1-Dichloroethene	0.644	0.588		-8.7	50
Acetone	0.294	0.351		19.39	50
Carbon Disulfide	1.767	1.404		-20.54	50
Methyl tert-butyl Ether	2.050	1.999		-2.49	50
Methylene Chloride	0.721	0.693		-3.88	50
trans-1,2-Dichloroethene	0.634	0.586		-7.57	50
1,1-Dichloroethane	1.233	1.217	0.1	-1.3	50
2-Butanone	0.478	0.582		21.76	50
Carbon Tetrachloride	0.460	0.456		-0.87	50
cis-1,2-Dichloroethene	0.762	0.735		-3.54	50
Chloroform	1.196	1.210		1.17	50
1,1,1-Trichloroethane	1.014	0.999		-1.48	50
Methylcyclohexane	0.606	0.586		-3.3	50
Benzene	1.465	1.440		-1.71	50
1,2-Dichloroethane	0.467	0.512		9.64	50
Trichloroethene	0.335	0.332		-0.9	50
1,2-Dichloropropane	0.364	0.369		1.37	50
Bromodichloromethane	0.489	0.527		7.77	50
4-Methyl-2-Pentanone	0.512	0.643		25.59	50
Toluene	0.872	0.881		1.03	50
t-1,3-Dichloropropene	0.495	0.493		-0.4	50
cis-1,3-Dichloropropene	0.552	0.553		0.18	50
1,1,2-Trichloroethane	0.335	0.357		6.57	50
2-Hexanone	0.369	0.477		29.27	50
Dibromochloromethane	0.359	0.388		8.08	50
Tetrachloroethene	0.315	0.301		-4.44	50
Chlorobenzene	1.074	1.075	0.3	0.09	50
Ethyl Benzene	1.895	1.889		-0.32	50
m/p-Xylenes	0.699	0.704		0.71	50
o-Xylene	0.701	0.697		-0.57	50
Styrene	1.130	1.186		4.96	50
Bromoform	0.258	0.284	0.1	10.08	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1401 SAS No.: Q1401 SDG No.: Q1401

Instrument ID: MSVOA_X Calibration Date/Time: 02/25/2025 16:28

Lab File ID: VX045042.D Init. Calib. Date(s): 02/10/2025 02/10/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:25 12:28

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Isopropylbenzene	4.026	3.925		-2.51	50
1,1,2,2-Tetrachloroethane	1.383	1.401	0.3	1.3	50
1,3-Dichlorobenzene	1.678	1.635		-2.56	50
1,4-Dichlorobenzene	1.697	1.651		-2.71	50
1,2-Dichlorobenzene	1.650	1.641		-0.55	50
1,2-Dichloroethane-d4	0.732	0.643		-12.16	50
Dibromofluoromethane	0.325	0.287		-11.69	50
Toluene-d8	1.229	1.046		-14.89	50
4-Bromofluorobenzene	0.414	0.377		-8.94	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.