

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022825\
 Data File : VX045073.D
 Acq On : 28 Feb 2025 03:47
 Operator : JC/MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC150

Quant Time: Feb 28 05:36:17 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 05:28:02 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	95754	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	176991	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	150547	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	67685	50.000	ug/l	# 0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	234624	165.865	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 331.740%#	
35) Dibromofluoromethane	5.379	113	179722	155.359	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 310.720%#	
50) Toluene-d8	8.647	98	638979	147.538	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 295.080%#	
62) 4-Bromofluorobenzene	11.079	95	223346	154.420	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 308.840%#	

Target Compounds				Qvalue		
2) Dichlorodifluoromethane	1.166	85	174297	132.067	ug/l	96
3) Chloromethane	1.307	50	213704	133.072	ug/l	97
4) Vinyl Chloride	1.374	62	217672	138.774	ug/l	97
5) Bromomethane	1.593	94	83547	170.379	ug/l	98
6) Chloroethane	1.660	64	82101	137.509	ug/l	95
7) Trichlorofluoromethane	1.861	101	272235	135.494	ug/l	98
8) Diethyl Ether	2.136	74	110005	150.660	ug/l	98
9) 1,1,2-Trichlorotrifluo...	2.313	101	161825	134.877	ug/l	99
10) Methyl Iodide	2.441	142	212355	138.059	ug/l	99
11) Tert butyl alcohol	3.014	59	195415	754.176	ug/l	97
12) 1,1-Dichloroethene	2.300	96	173271	140.655	ug/l	96
13) Acrolein	2.239	56	263999	940.846	ug/l	98
14) Allyl chloride	2.654	41	308018	139.077	ug/l	98
15) Acrylonitrile	3.069	53	573208	813.990	ug/l	99
16) Acetone	2.392	43	511828	879.387	ug/l	100
17) Carbon Disulfide	2.495	76	487720	145.511	ug/l	98
18) Methyl Acetate	2.709	43	258303	147.347	ug/l	100
19) Methyl tert-butyl Ether	3.117	73	619778	157.358	ug/l	99
20) Methylene Chloride	2.782	84	202913	147.198	ug/l	99
21) trans-1,2-Dichloroethene	3.081	96	176924	146.027	ug/l	98
22) Diisopropyl ether	3.764	45	662828	156.526	ug/l	98
23) Vinyl Acetate	3.721	43	2943150	846.586	ug/l	99
24) 1,1-Dichloroethane	3.605	63	363081	153.453	ug/l	99
25) 2-Butanone	4.568	43	818497	872.239	ug/l	99
26) 2,2-Dichloropropane	4.465	77	179826	102.231	ug/l	99
27) cis-1,2-Dichloroethene	4.483	96	220826	151.205	ug/l	98
28) Bromochloromethane	4.891	49	171273	150.105	ug/l	100
29) Tetrahydrofuran	5.007	42	528574	850.174	ug/l	99
30) Chloroform	5.093	83	351934	152.067	ug/l	97
31) Cyclohexane	5.458	56	302215	139.174	ug/l	96
32) 1,1,1-Trichloroethane	5.379	97	294434	151.120	ug/l	100
36) 1,1-Dichloropropene	5.684	75	241405	145.112	ug/l	100
37) Ethyl Acetate	4.721	43	316250	167.067	ug/l	99
38) Carbon Tetrachloride	5.672	117	245917	150.363	ug/l	99
39) Methylcyclohexane	7.373	83	292216	137.544	ug/l	97
40) Benzene	6.031	78	755854	145.368	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022825\
 Data File : VX045073.D
 Acq On : 28 Feb 2025 03:47
 Operator : JC/MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC150

Quant Time: Feb 28 05:36:17 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 05:28:02 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.928	41	175023	173.088	ug/l	99
42) 1,2-Dichloroethane	6.080	62	276129	163.872	ug/l	99
43) Isopropyl Acetate	6.342	43	511643	169.804	ug/l	99
44) Trichloroethene	7.123	130	178483	150.181	ug/l	97
45) 1,2-Dichloropropane	7.428	63	198135	153.334	ug/l	100
46) Dibromomethane	7.580	93	142119	157.443	ug/l	99
47) Bromodichloromethane	7.818	83	280255	160.492	ug/l	99
48) Methyl methacrylate	7.696	41	252275	176.370	ug/l	99
49) 1,4-Dioxane	7.665	88	96268	3166.423	ug/l	99
51) 4-Methyl-2-Pentanone	8.580	43	1536157	832.051	ug/l	99
52) Toluene	8.714	92	448664	145.321	ug/l	99
53) t-1,3-Dichloropropene	8.976	75	266444	154.373	ug/l	96
54) cis-1,3-Dichloropropene	8.366	75	293521	152.117	ug/l	96
55) 1,1,2-Trichloroethane	9.147	97	178217	149.046	ug/l	98
56) Ethyl methacrylate	9.116	69	310992	162.631	ug/l	98
57) 1,3-Dichloropropane	9.305	76	312196	149.699	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.238	63	717328	764.019	ug/l	99
59) 2-Hexanone	9.433	43	1156653	868.492	ug/l	100
60) Dibromochloromethane	9.519	129	201906	158.276	ug/l	99
61) 1,2-Dibromoethane	9.610	107	185828	154.074	ug/l	100
64) Tetrachloroethene	9.269	164	139536	145.906	ug/l	94
65) Chlorobenzene	10.079	112	472117	146.096	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	163370	156.845	ug/l	97
67) Ethyl Benzene	10.189	91	852575	149.191	ug/l	99
68) m/p-Xylenes	10.299	106	607523	288.655	ug/l	97
69) o-Xylene	10.640	106	302495	143.259	ug/l	99
70) Styrene	10.653	104	512379	150.979	ug/l	99
71) Bromoform	10.799	173	135388	174.808	ug/l #	99
73) Isopropylbenzene	10.963	105	780706	142.719	ug/l	99
74) N-amyl acetate	10.842	43	429686	173.042	ug/l	100
75) 1,1,2,2-Tetrachloroethane	11.213	83	283558	150.458	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	303633	201.103	ug/l	86
77) Bromobenzene	11.195	156	180997	145.760	ug/l	96
78) n-propylbenzene	11.305	91	923593	146.395	ug/l	99
79) 2-Chlorotoluene	11.366	91	548695	141.695	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	635832	143.293	ug/l	98
81) trans-1,4-Dichloro-2-b...	11.018	75	82267	147.761	ug/l	98
82) 4-Chlorotoluene	11.451	91	627816	146.464	ug/l	99
83) tert-Butylbenzene	11.713	119	641713	142.974	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	647248	148.135	ug/l	100
85) sec-Butylbenzene	11.890	105	807713	145.383	ug/l	100
86) p-Isopropyltoluene	12.006	119	664916	149.075	ug/l	100
87) 1,3-Dichlorobenzene	11.969	146	334751	147.488	ug/l	100
88) 1,4-Dichlorobenzene	12.043	146	333594	145.189	ug/l	99
89) n-Butylbenzene	12.329	91	620757	158.324	ug/l	99
90) Hexachloroethane	12.536	117	128582	154.554	ug/l	100
91) 1,2-Dichlorobenzene	12.335	146	329318	147.436	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	64958	186.240	ug/l	98
93) 1,2,4-Trichlorobenzene	13.585	180	219313	162.416	ug/l	99
94) Hexachlorobutadiene	13.725	225	80241	149.034	ug/l	99
95) Naphthalene	13.774	128	828949	168.538	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	222608	162.727	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022825\
Data File : VX045073.D
Acq On : 28 Feb 2025 03:47
Operator : JC/MD
Sample : VSTDICC150
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Quant Time: Feb 28 05:36:17 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 05:28:02 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022825\
 Data File : VX045073.D
 Acq On : 28 Feb 2025 03:47
 Operator : JC/MD
 Sample : VSTDIC150
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC150

Quant Time: Feb 28 05:36:17 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 05:28:02 2025
 Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion:168 Resp: 95754
Ion Ratio Lower Upper
168 100
99 64.8 48.2 72.4



#2
Dichlorodifluoromethane
Concen: 132.067 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47

Tgt Ion: 85 Resp: 174297
Ion Ratio Lower Upper
85 100
87 32.5 17.4 52.3









#3
Chloromethane
Concen: 133.072 ug/l
RT: 1.307 min Scan# 30
Delta R.T. -0.000 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 50 Resp: 213704
Ion Ratio Lower Upper
50 100
52 32.7 25.0 37.4



#4
Vinyl Chloride
Concen: 138.774 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47

Tgt Ion: 62 Resp: 217672
Ion Ratio Lower Upper
62 100
64 31.8 24.3 36.5









#5
Bromomethane
Concen: 170.379 ug/l
RT: 1.593 min Scan# 81
Delta R.T. -0.000 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 94 Resp: 83547
Ion Ratio Lower Upper
94 100
96 92.2 75.0 112.4



#6
Chloroethane
Concen: 137.509 ug/l
RT: 1.660 min Scan# 94
Delta R.T. -0.000 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47

Tgt Ion: 64 Resp: 82101
Ion Ratio Lower Upper
64 100
66 30.8 26.7 40.1









#7
Trichlorofluoromethane
Concen: 135.494 ug/l
RT: 1.861 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion:101 Resp: 272235
Ion Ratio Lower Upper
101 100
103 66.4 52.1 78.1



#8
Diethyl Ether
Concen: 150.660 ug/l
RT: 2.136 min Scan# 172
Delta R.T. -0.000 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47

Tgt Ion: 74 Resp: 110005
Ion Ratio Lower Upper
74 100
45 104.9 51.5 154.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 134.877 ug/l
RT: 2.313 min Scan# 201
Delta R.T. -0.000 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion:101 Resp: 161825
Ion Ratio Lower Upper
101 100
85 45.8 36.2 54.4
151 70.8 56.4 84.6



#10
Methyl Iodide
Concen: 138.059 ug/l
RT: 2.441 min Scan# 222
Delta R.T. -0.000 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47

Tgt Ion:142 Resp: 212355
Ion Ratio Lower Upper
142 100
127 45.2 35.4 53.2
141 14.3 11.6 17.4









#11
Tert butyl alcohol
Concen: 754.176 ug/l
RT: 3.014 min Scan# 311
Delta R.T. 0.030 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 59 Resp: 195415
Ion Ratio Lower Upper
59 100
57 10.9 7.8 11.8



#12
1,1-Dichloroethene
Concen: 140.655 ug/l
RT: 2.300 min Scan# 199
Delta R.T. -0.006 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47

Tgt Ion: 96 Resp: 173271
Ion Ratio Lower Upper
96 100
61 174.5 134.6 202.0
98 65.6 51.0 76.6





#13

Acrolein

Concen: 940.846 ug/l

RT: 2.239 min Scan# 188

Delta R.T. 0.006 min

Lab File: VX045073.D

Acq: 28 Feb 2025 03:47

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC150

Tgt Ion: 56 Resp: 263999

Ion Ratio Lower Upper

56 100

55 71.6 56.2 84.4



#14

Allyl chloride

Concen: 139.077 ug/l

RT: 2.654 min Scan# 257

Delta R.T. -0.000 min

Lab File: VX045073.D

Acq: 28 Feb 2025 03:47




Tgt Ion: 41 Resp: 308018


Ion Ratio Lower Upper

41 100

39 67.9 53.8 80.8

76 33.8 25.2 37.8









#15
Acrylonitrile
Concen: 813.990 ug/l
RT: 3.069 min Scan# 31
Delta R.T. 0.006 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 53 Resp: 573208
Ion Ratio Lower Upper
53 100
52 83.3 66.2 99.2
51 37.0 29.0 43.4



#16
Acetone
Concen: 879.387 ug/l
RT: 2.392 min Scan# 214
Delta R.T. 0.006 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47

Tgt Ion: 43 Resp: 511828
Ion Ratio Lower Upper
43 100
58 30.3 24.2 36.4





#17

Carbon Disulfide

Concen: 145.511 ug/l

RT: 2.495 min Scan# 21

Delta R.T. -0.006 min

Lab File: VX045073.D

Acq: 28 Feb 2025 03:47

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC150

Tgt Ion: 76 Resp: 487720

Ion Ratio Lower Upper

76 100

78 8.8 6.6 9.8



#18

Methyl Acetate

Concen: 147.347 ug/l

RT: 2.709 min Scan# 266

Delta R.T. 0.006 min

Lab File: VX045073.D




Acq: 28 Feb 2025 03:47


Tgt Ion: 43 Resp: 258303

Ion Ratio Lower Upper

43 100

74 22.0 17.4 26.2









#19
Methyl tert-butyl Ether
Concen: 157.358 ug/l
RT: 3.117 min Scan# 311
Delta R.T. -0.000 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 73 Resp: 619778
Ion Ratio Lower Upper
73 100
57 22.4 18.5 27.7



#20
Methylene Chloride
Concen: 147.198 ug/l
RT: 2.782 min Scan# 278
Delta R.T. -0.000 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47

Tgt Ion: 84 Resp: 202913
Ion Ratio Lower Upper
84 100
49 133.6 106.5 159.7
51 40.9 32.1 48.1
86 64.8 50.0 75.0









#21
trans-1,2-Dichloroethene
Concen: 146.027 ug/l
RT: 3.081 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 96 Resp: 176924
Ion Ratio Lower Upper
96 100
61 147.0 117.0 175.4
98 63.6 53.4 80.2



#22
Diisopropyl ether
Concen: 156.526 ug/l
RT: 3.764 min Scan# 439
Delta R.T. -0.000 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47

Tgt Ion: 45 Resp: 662828
Ion Ratio Lower Upper
45 100
43 70.0 54.9 82.3
87 25.2 21.0 31.4
59 10.9 8.7 13.1









#23
 Vinyl Acetate
 Concen: 846.586 ug/l
 RT: 3.721 min Scan# 413
 Delta R.T. -0.000 min
 Lab File: VX045073.D
 Acq: 28 Feb 2025 03:47




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC150


Tgt Ion: 43 Resp: 2943150
 Ion Ratio Lower Upper
 43 100
 86 9.8 8.1 12.1



#24
 1,1-Dichloroethane
 Concen: 153.453 ug/l
 RT: 3.605 min Scan# 413
 Delta R.T. -0.000 min
 Lab File: VX045073.D
 Acq: 28 Feb 2025 03:47

Tgt Ion: 63 Resp: 363081
 Ion Ratio Lower Upper
 63 100
 98 7.1 3.4 10.2
 100 4.4 2.1 6.5









#25
2-Butanone
Concen: 872.239 ug/l
RT: 4.568 min Scan# 51
Delta R.T. 0.006 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 43 Resp: 818497
Ion Ratio Lower Upper
43 100
72 24.7 20.0 30.0



#26
2,2-Dichloropropane
Concen: 102.231 ug/l
RT: 4.465 min Scan# 554
Delta R.T. -0.006 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47

Tgt Ion: 77 Resp: 179826
Ion Ratio Lower Upper
77 100
97 24.1 12.4 37.0





#27

cis-1,2-Dichloroethene

Concen: 151.205 ug/l

RT: 4.483 min Scan# 51

Delta R.T. -0.000 min

Lab File: VX045073.D

Acq: 28 Feb 2025 03:47

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150





Tgt Ion: 96 Resp: 220826

Ion Ratio Lower Upper

96 100

61 144.2 0.0 283.2

98 64.3 0.0 128.0



#28

Bromochloromethane

Concen: 150.105 ug/l

RT: 4.891 min Scan# 624

Delta R.T. -0.006 min

Lab File: VX045073.D

Acq: 28 Feb 2025 03:47




Tgt Ion: 49 Resp: 171273


Ion Ratio Lower Upper

49 100

129 1.7 0.0 3.4

130 70.3 56.1 84.1





#29

Tetrahydrofuran

Concen: 850.174 ug/l

RT: 5.007 min Scan# 64

Delta R.T. -0.000 min

Lab File: VX045073.D

Acq: 28 Feb 2025 03:47

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150





Tgt Ion: 42 Resp: 528574

Ion Ratio Lower Upper

42 100

72 42.8 34.1 51.1

71 39.7 31.4 47.0



#30

Chloroform

Concen: 152.067 ug/l

RT: 5.093 min Scan# 657

Delta R.T. 0.006 min

Lab File: VX045073.D




Acq: 28 Feb 2025 03:47


Tgt Ion: 83 Resp: 351934

Ion Ratio Lower Upper

83 100

85 66.9 51.4 77.2









#31
Cyclohexane
Concen: 139.174 ug/l
RT: 5.458 min Scan# 717
Delta R.T. -0.006 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 56 Resp: 302215
Ion Ratio Lower Upper
56 100
69 28.8 23.4 35.2
84 82.2 69.4 104.2



#32
1,1,1-Trichloroethane
Concen: 151.120 ug/l
RT: 5.379 min Scan# 704
Delta R.T. -0.000 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47

Tgt Ion: 97 Resp: 294434
Ion Ratio Lower Upper
97 100
99 64.7 51.6 77.4
61 48.1 38.4 57.6





#33

1,2-Dichloroethane-d4

Concen: 165.865 ug/l

RT: 5.952 min Scan# 798

Delta R.T. -0.000 min

Lab File: VX045073.D

Acq: 28 Feb 2025 03:47

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC150

Tgt Ion: 65 Resp: 234624

Ion Ratio Lower Upper

65 100

67 52.7 0.0 106.2



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. -0.000 min

Lab File: VX045073.D

Acq: 28 Feb 2025 03:47




Tgt Ion: 114 Resp: 176991


Ion Ratio Lower Upper

114 100

63 22.8 0.0 41.8

88 16.4 0.0 32.8





#35

Dibromofluoromethane

Concen: 155.359 ug/l

RT: 5.379 min Scan# 705

Delta R.T. -0.006 min

Lab File: VX045073.D

Acq: 28 Feb 2025 03:47

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150





Tgt Ion:113 Resp: 179722

Ion Ratio Lower Upper

113 100

111 102.7 81.8 122.6

192 17.5 14.3 21.5



#36

1,1-Dichloropropene

Concen: 145.112 ug/l

RT: 5.684 min Scan# 754

Delta R.T. -0.006 min

Lab File: VX045073.D

Acq: 28 Feb 2025 03:47




Tgt Ion: 75 Resp: 241405


Ion Ratio Lower Upper

75 100

110 33.7 16.9 50.6

77 30.5 24.5 36.7





#37

Ethyl Acetate

Concen: 167.067 ug/l

RT: 4.721 min Scan# 596

Delta R.T. 0.006 min

Lab File: VX045073.D

Acq: 28 Feb 2025 03:47

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150





Tgt Ion: 43 Resp: 316250

Ion Ratio Lower Upper

43 100

61 13.2 10.8 16.2

70 10.1 7.7 11.5



#38

Carbon Tetrachloride

Concen: 150.363 ug/l

RT: 5.672 min Scan# 752

Delta R.T. -0.000 min

Lab File: VX045073.D

Acq: 28 Feb 2025 03:47




Tgt Ion: 117 Resp: 245917


Ion Ratio Lower Upper

117 100

119 95.9 76.7 115.1

121 30.7 25.5 38.3









#39
Methylcyclohexane
Concen: 137.544 ug/l
RT: 7.373 min Scan# 1031
Delta R.T. -0.000 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 83 Resp: 292216
Ion Ratio Lower Upper
83 100
55 81.7 63.0 94.4
98 47.4 39.7 59.5



#40
Benzene
Concen: 145.368 ug/l
RT: 6.031 min Scan# 811
Delta R.T. -0.000 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47

Tgt Ion: 78 Resp: 755854
Ion Ratio Lower Upper
78 100
77 23.7 18.8 28.2









#41
Methacrylonitrile
Concen: 173.088 ug/l
RT: 4.928 min Scan# 61
Delta R.T. 0.006 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150

Tgt Ion: 41 Resp: 175023




Ion	Ratio	Lower	Upper
41	100		
39	55.9	44.6	67.0
67	69.5	53.9	80.9
52	29.6	23.3	34.9




#42
1,2-Dichloroethane
Concen: 163.872 ug/l
RT: 6.080 min Scan# 819
Delta R.T. -0.000 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47

Tgt Ion: 62 Resp: 276129

Ion	Ratio	Lower	Upper
62	100		
98	8.8	0.0	18.2









#43
Isopropyl Acetate
Concen: 169.804 ug/l
RT: 6.342 min Scan# 862
Delta R.T. -0.000 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 43 Resp: 511643
Ion Ratio Lower Upper
43 100
61 18.7 14.9 22.3
87 12.1 9.4 14.2



#44
Trichloroethene
Concen: 150.181 ug/l
RT: 7.123 min Scan# 990
Delta R.T. -0.000 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47

Tgt Ion:130 Resp: 178483
Ion Ratio Lower Upper
130 100
95 99.3 0.0 205.0





#45

1,2-Dichloropropane

Concen: 153.334 ug/l

RT: 7.428 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VX045073.D

Acq: 28 Feb 2025 03:47

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC150

Tgt Ion: 63 Resp: 198135

Ion Ratio Lower Upper

63 100

65 30.8 24.7 37.1



#46

Dibromomethane

Concen: 157.443 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. -0.000 min

Lab File: VX045073.D

Acq: 28 Feb 2025 03:47




Tgt Ion: 93 Resp: 142119


Ion Ratio Lower Upper

93 100

95 82.6 65.8 98.8

174 88.2 72.2 108.2









#47
Bromodichloromethane
Concen: 160.492 ug/l
RT: 7.818 min Scan# 1104
Delta R.T. -0.000 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 83 Resp: 280255
Ion Ratio Lower Upper
83 100
85 64.9 51.1 76.7
127 7.8 6.4 9.6



#48
Methyl methacrylate
Concen: 176.370 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. 0.006 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47

Tgt Ion: 41 Resp: 252275
Ion Ratio Lower Upper
41 100
69 78.1 63.0 94.6
39 61.1 47.5 71.3





#49

1,4-Dioxane

Concen: 3166.423 ug/l

RT: 7.665 min Scan# 1078

Delta R.T. 0.006 min

Lab File: VX045073.D

Acq: 28 Feb 2025 03:47

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150





Tgt Ion: 88 Resp: 96268

Ion Ratio Lower Upper

88 100

43 37.0 28.7 43.1

58 70.8 55.8 83.8



#50

Toluene-d8

Concen: 147.538 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. -0.000 min

Lab File: VX045073.D




Acq: 28 Feb 2025 03:47


Tgt Ion: 98 Resp: 638979

Ion Ratio Lower Upper

98 100

100 65.4 52.0 78.0









#51
4-Methyl-2-Pentanone
Concen: 832.051 ug/l
RT: 8.580 min Scan# 1229
Delta R.T. 0.006 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150


Tgt Ion: 43 Resp: 1536157
Ion Ratio Lower Upper
43 100
58 37.1 29.2 43.8



#52
Toluene
Concen: 145.321 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. -0.000 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47

Tgt Ion: 92 Resp: 448664
Ion Ratio Lower Upper
92 100
91 172.3 138.9 208.3





#53

t-1,3-Dichloropropene

Concen: 154.373 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX045073.D

Acq: 28 Feb 2025 03:47

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC150

Tgt Ion: 75 Resp: 266444

Ion Ratio Lower Upper

75 100

77 30.9 26.5 39.7



#54

cis-1,3-Dichloropropene

Concen: 152.117 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX045073.D

Acq: 28 Feb 2025 03:47




Tgt Ion: 75 Resp: 293521


Ion Ratio Lower Upper

75 100

77 32.3 24.7 37.1

39 54.0 40.7 61.1









#55
1,1,2-Trichloroethane
Concen: 149.046 ug/l
RT: 9.147 min Scan# 1322
Delta R.T. -0.000 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150

Tgt Ion: 97 Resp: 178217




Ion	Ratio	Lower	Upper
97	100		
83	87.8	69.4	104.2
85	57.8	44.1	66.1
99	63.1	50.7	76.1




#56
Ethyl methacrylate
Concen: 162.631 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. -0.000 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47

Tgt Ion: 69 Resp: 310992

Ion	Ratio	Lower	Upper
69	100		
41	72.6	57.0	85.4
39	44.2	34.2	51.4





#57

1,3-Dichloropropane

Concen: 149.699 ug/l

RT: 9.305 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX045073.D

Acq: 28 Feb 2025 03:47

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC150

Tgt Ion: 76 Resp: 312196

Ion Ratio Lower Upper

76 100

78 32.0 25.5 38.3



#58

2-Chloroethyl Vinyl ether

Concen: 764.019 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. -0.000 min

Lab File: VX045073.D




Acq: 28 Feb 2025 03:47


Tgt Ion: 63 Resp: 717328

Ion Ratio Lower Upper

63 100

106 26.5 21.5 32.3





#59

2-Hexanone

Concen: 868.492 ug/l

RT: 9.433 min Scan# 1369

Delta R.T. 0.006 min

Lab File: VX045073.D

Acq: 28 Feb 2025 03:47

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC150

Tgt Ion: 43 Resp: 1156653

Ion Ratio Lower Upper

43 100

58 51.6 25.9 77.6



#60

Dibromochloromethane

Concen: 158.276 ug/l

RT: 9.519 min Scan# 1383

Delta R.T. -0.000 min

Lab File: VX045073.D




Acq: 28 Feb 2025 03:47


Tgt Ion:129 Resp: 201906

Ion Ratio Lower Upper

129 100

127 77.9 38.5 115.5









#61
1,2-Dibromoethane
Concen: 154.074 ug/l
RT: 9.610 min Scan# 1397
Delta R.T. 0.006 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion:107 Resp: 185828
Ion Ratio Lower Upper
107 100
109 95.2 76.1 114.1



#62
4-Bromofluorobenzene
Concen: 154.420 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47

Tgt Ion: 95 Resp: 223346
Ion Ratio Lower Upper
95 100
174 70.4 0.0 148.2
176 69.4 0.0 141.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.006 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion:117 Resp: 150547
Ion Ratio Lower Upper
117 100
82 56.4 46.3 69.5
119 32.1 25.7 38.5



#64
Tetrachloroethene
Concen: 145.906 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. -0.000 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47

Tgt Ion:164 Resp: 139536
Ion Ratio Lower Upper
164 100
166 121.6 103.6 155.4
129 99.0 82.7 124.1
131 93.6 80.5 120.7









#65
Chlorobenzene
Concen: 146.096 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion:112 Resp: 472117
Ion Ratio Lower Upper
112 100
114 31.9 26.0 39.0



#66
1,1,1,2-Tetrachloroethane
Concen: 156.845 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47

Tgt Ion:131 Resp: 163370
Ion Ratio Lower Upper
131 100
133 94.2 48.6 145.9
119 66.0 33.9 101.7









#67
Ethyl Benzene
Concen: 149.191 ug/l
RT: 10.189 min Scan# 1493
Delta R.T. -0.000 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150


Tgt Ion: 91 Resp: 852575
Ion Ratio Lower Upper
91 100
106 29.5 23.9 35.9



#68
m/p-Xylenes
Concen: 288.655 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47

Tgt Ion:106 Resp: 607523
Ion Ratio Lower Upper
106 100
91 211.2 165.4 248.0









#69
o-Xylene
Concen: 143.259 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion:106 Resp: 302495
Ion Ratio Lower Upper
106 100
91 221.3 109.9 329.6



#70
Styrene
Concen: 150.979 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47

Tgt Ion:104 Resp: 512379
Ion Ratio Lower Upper
104 100
78 54.4 42.2 63.4
103 54.8 43.8 65.8





#71

Bromoform

Concen: 174.808 ug/l

RT: 10.799 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX045073.D

Acq: 28 Feb 2025 03:47

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150





Tgt Ion:173 Resp: 135388

Ion Ratio Lower Upper

173 100

175 48.3 24.6 73.6

254 0.1 0.0 0.0#



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.024 min Scan# 1794

Delta R.T. 0.006 min

Lab File: VX045073.D

Acq: 28 Feb 2025 03:47




Tgt Ion:152 Resp: 67685


Ion Ratio Lower Upper

152 100

115 138.4 44.2 132.6#

150 208.8 0.0 349.0





#73

Isopropylbenzene

Concen: 142.719 ug/l

RT: 10.963 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX045073.D

Acq: 28 Feb 2025 03:47

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC150

Tgt Ion:105 Resp: 780706

Ion Ratio Lower Upper

105 100

120 26.0 13.2 39.5



#74

N-ethyl acetate

Concen: 173.042 ug/l

RT: 10.842 min Scan# 1600

Delta R.T. -0.000 min

Lab File: VX045073.D

Acq: 28 Feb 2025 03:47

Tgt Ion: 43 Resp: 429686




Ion Ratio Lower Upper


43 100

70 39.6 31.8 47.6

55 22.5 18.3 27.5

61 23.1 18.3 27.5





#75

1,1,2,2-Tetrachloroethane

Concen: 150.458 ug/l

RT: 11.213 min Scan# 1661

Delta R.T. -0.000 min

Lab File: VX045073.D

Acq: 28 Feb 2025 03:47

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150





Tgt Ion: 83 Resp: 283558

Ion Ratio Lower Upper

83 100

131 9.3 4.5 13.4

85 63.8 31.7 95.1



#76

1,2,3-Trichloropropane

Concen: 201.103 ug/l

RT: 11.238 min Scan# 1665

Delta R.T. -0.000 min

Lab File: VX045073.D




Acq: 28 Feb 2025 03:47


Tgt Ion: 75 Resp: 303633

Ion Ratio Lower Upper

75 100

77 32.6 20.7 62.1





#77

Bromobenzene

Concen: 145.760 ug/l

RT: 11.195 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX045073.D

Acq: 28 Feb 2025 03:47

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150





Tgt Ion:156 Resp: 180997

Ion Ratio Lower Upper

156 100

77 179.4 85.8 257.4

158 97.1 48.4 145.2



#78

n-propylbenzene

Concen: 146.395 ug/l

RT: 11.305 min Scan# 1676

Delta R.T. -0.000 min

Lab File: VX045073.D




Acq: 28 Feb 2025 03:47


Tgt Ion: 91 Resp: 923593

Ion Ratio Lower Upper

91 100

120 21.8 11.2 33.6





#79

2-Chlorotoluene

Concen: 141.695 ug/l

RT: 11.366 min Scan# 1

Delta R.T. 0.006 min

Lab File: VX045073.D

Acq: 28 Feb 2025 03:47

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC150

Tgt Ion: 91 Resp: 548695

Ion Ratio Lower Upper

91 100

126 32.4 16.4 49.4



#80

1,3,5-Trimethylbenzene

Concen: 143.293 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. -0.000 min

Lab File: VX045073.D




Acq: 28 Feb 2025 03:47


Tgt Ion: 105 Resp: 635832

Ion Ratio Lower Upper

105 100

120 46.7 24.1 72.2









#81
trans-1,4-Dichloro-2-butene
Concen: 147.761 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.000 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 75 Resp: 82267
Ion Ratio Lower Upper
75 100
53 107.7 88.5 132.7
89 45.2 36.2 54.4



#82
4-Chlorotoluene
Concen: 146.464 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47

Tgt Ion: 91 Resp: 627816
Ion Ratio Lower Upper
91 100
126 27.8 14.1 42.4





#83

tert-Butylbenzene

Concen: 142.974 ug/l

RT: 11.713 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX045073.D

Acq: 28 Feb 2025 03:47

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC150





Tgt Ion:119 Resp: 641713

Ion Ratio Lower Upper

119 100

91 62.5 30.7 92.1

134 23.0 11.7 35.0



#84

1,2,4-Trimethylbenzene

Concen: 148.135 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. -0.000 min

Lab File: VX045073.D




Acq: 28 Feb 2025 03:47


Tgt Ion:105 Resp: 647248

Ion Ratio Lower Upper

105 100

120 43.8 22.1 66.1









#85
 sec-Butylbenzene
 Concen: 145.383 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. -0.000 min
 Lab File: VX045073.D
 Acq: 28 Feb 2025 03:47




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC150


Tgt Ion:105 Resp: 807713
 Ion Ratio Lower Upper
 105 100
 134 19.4 9.8 29.4



#86
 p-Isopropyltoluene
 Concen: 149.075 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. -0.000 min
 Lab File: VX045073.D
 Acq: 28 Feb 2025 03:47

Tgt Ion:119 Resp: 664916
 Ion Ratio Lower Upper
 119 100
 134 25.7 12.9 38.6
 91 25.8 12.7 38.0









#87
1,3-Dichlorobenzene
Concen: 147.488 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. -0.000 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion:146 Resp: 334751
Ion Ratio Lower Upper
146 100
111 42.7 21.1 63.4
148 63.6 31.9 95.7



#88
1,4-Dichlorobenzene
Concen: 145.189 ug/l
RT: 12.043 min Scan# 1797
Delta R.T. 0.006 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47

Tgt Ion:146 Resp: 333594
Ion Ratio Lower Upper
146 100
111 41.7 20.2 60.5
148 64.0 31.9 95.9





#89

n-Butylbenzene

Concen: 158.324 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. -0.000 min

Lab File: VX045073.D

Acq: 28 Feb 2025 03:47

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150





Tgt Ion: 91 Resp: 620757

Ion Ratio Lower Upper

91 100

92 53.9 26.7 80.0

134 23.6 12.2 36.6



#90

Hexachloroethane

Concen: 154.554 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. -0.000 min

Lab File: VX045073.D




Acq: 28 Feb 2025 03:47


Tgt Ion: 117 Resp: 128582

Ion Ratio Lower Upper

117 100

201 64.2 31.9 95.9









#91
1,2-Dichlorobenzene
Concen: 147.436 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion:146 Resp: 329318
Ion Ratio Lower Upper
146 100
111 44.9 21.6 65.0
148 63.4 31.9 95.9



#92
1,2-Dibromo-3-Chloropropane
Concen: 186.240 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. -0.000 min
Lab File: VX045073.D
Acq: 28 Feb 2025 03:47

Tgt Ion: 75 Resp: 64958
Ion Ratio Lower Upper
75 100
155 77.9 39.6 118.7
157 100.1 51.1 153.4





#93

1,2,4-Trichlorobenzene

Concen: 162.416 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. -0.000 min

Lab File: VX045073.D

Acq: 28 Feb 2025 03:47

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150





Tgt Ion:180 Resp: 219313

Ion Ratio Lower Upper

180 100

182 95.6 47.3 141.9

145 32.9 16.3 48.9



#94

Hexachlorobutadiene

Concen: 149.034 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. -0.000 min

Lab File: VX045073.D

Acq: 28 Feb 2025 03:47




Tgt Ion:225 Resp: 80241


Ion Ratio Lower Upper

225 100

223 63.2 31.9 95.7

227 63.4 31.4 94.0





#95

Naphthalene

Concen: 168.538 ug/l

RT: 13.774 min Scan# 2081

Delta R.T. -0.000 min

Lab File: VX045073.D

Acq: 28 Feb 2025 03:47

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150





Tgt Ion:128 Resp: 828949

Ion Ratio Lower Upper

128 100

127 12.8 10.3 15.5

129 11.1 8.7 13.1



#96

1,2,3-Trichlorobenzene

Concen: 162.727 ug/l

RT: 13.957 min Scan# 2111

Delta R.T. -0.000 min

Lab File: VX045073.D

Acq: 28 Feb 2025 03:47

Tgt Ion:180 Resp: 222608

Ion Ratio Lower Upper

180 100

182 95.7 47.5 142.6

145 35.5 17.0 50.9

