

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q1401 SAS No.: Q1401 SDG NO.: Q1401
Lab File ID: VX045077.D Date Analyzed: 02/28/2025
Instrument ID: MSVOA_X Time Analyzed: 10:32
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	111989	5.54	196414	6.75	172369	10.05
UPPER LIMIT	223978	6.037	392828	7.251	344738	10.549
LOWER LIMIT	55994.5	5.037	98207	6.251	86184.5	9.549
EPA SAMPLE NO.						
BP-VPB-192-GW-840-842	67792	5.54	135879	6.76	123593	10.05
BP-VPB-192-GW-900-902	73969	5.55	147122	6.76	132699	10.06
VX0228WBL01	76430	5.54	149381	6.76	137180	10.05
VX0228WBS01	108329	5.54	192724	6.76	169474	10.05

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.