

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q1401 SAS No.: Q1401 SDG NO.: Q1401
Lab File ID: VY021320.D Date Analyzed: 02/26/2025
Instrument ID: MSVOA_Y Time Analyzed: 10:11
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	193888	7.71	308714	8.62	268805	11.42
UPPER LIMIT	387776	8.213	617428	9.116	537610	11.92
LOWER LIMIT	96944	7.213	154357	8.116	134403	10.92
EPA SAMPLE NO.						
BP-VPB-192-GW-825-827	237643	7.71	467253	8.62	479781	11.41
BP-VPB-192-GW-860-862	234926	7.71	466686	8.62	492462	11.41
BP-VPB-192-GW-880-882	240939	7.71	470542	8.61	479899	11.41
VY0226SBL01	251129	7.71	483695	8.62	470141	11.42
VY0226SBS01	184417	7.71	291138	8.62	254891	11.41
VY0226SBSD01	186200	7.71	292071	8.62	258744	11.41

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.