

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q1401 SAS No.: Q1401 SDG NO.: Q1401
Lab File ID: VX045076.D BFB Injection Date: 02/28/2025
Instrument ID: MSVOA_X BFB Injection Time: 10:03
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	21.1
75	30.0 - 60.0% of mass 95	52.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.9
173	Less than 2.0% of mass 174	0.8 (1) 1
174	50.0 - 100.0% of mass 95	74.6
175	5.0 - 9.0% of mass 174	5.5 (7.4) 1
176	95.0 - 101.0% of mass 174	72.8 (97.6) 1
177	5.0 - 9.0% of mass 176	5.1 (7) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX045077.D	02/28/2025	10:32
VX0228WBL01	VX0228WBL01	VX045079.D	02/28/2025	11:23
VX0228WBS01	VX0228WBS01	VX045080.D	02/28/2025	11:46
BP-VPB-192-GW-840-842	Q1401-03	VX045082.D	02/28/2025	12:37
BP-VPB-192-GW-900-902	Q1401-06	VX045083.D	02/28/2025	13:00
VSTDCCC050EC	VSTDCCC050	VX045098.D	02/28/2025	18:50