

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK  
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: TETR06  
Lab Code: CHEM Case No.: Q1401 SAS No.: Q1401 SDG NO.: Q1401  
Lab File ID: VY021319.D BFB Injection Date: 02/26/2025  
Instrument ID: MSVOA\_Y BFB Injection Time: 09:12  
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N Y

| m/e | ION ABUNDANCE CRITERIA             | % RELATIVE ABUNDANCE |
|-----|------------------------------------|----------------------|
| 50  | 15.0 - 40.0% of mass 95            | 21.5                 |
| 75  | 30.0 - 60.0% of mass 95            | 55.4                 |
| 95  | Base Peak, 100% relative abundance | 100                  |
| 96  | 5.0 - 9.0% of mass 95              | 6.6                  |
| 173 | Less than 2.0% of mass 174         | 1.2 ( 1.5 ) 1        |
| 174 | 50.0 - 100.0% of mass 95           | 78.5                 |
| 175 | 5.0 - 9.0% of mass 174             | 6.5 ( 8.3 ) 1        |
| 176 | 95.0 - 101.0% of mass 174          | 75.9 ( 96.7 ) 1      |
| 177 | 5.0 - 9.0% of mass 176             | 5.1 ( 6.7 ) 2        |

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

| EPA<br>SAMPLE NO.     | LAB<br>SAMPLE ID | LAB<br>FILE ID | DATE<br>ANALYZED | TIME<br>ANALYZED |
|-----------------------|------------------|----------------|------------------|------------------|
| VSTDCCC050            | VSTDCCC050       | VY021320.D     | 02/26/2025       | 10:11            |
| VY0226SBL01           | VY0226SBL01      | VY021321.D     | 02/26/2025       | 10:50            |
| VY0226SBS01           | VY0226SBS01      | VY021322.D     | 02/26/2025       | 11:22            |
| VY0226SBSD01          | VY0226SBSD01     | VY021323.D     | 02/26/2025       | 11:44            |
| BP-VPB-192-GW-825-827 | Q1401-07         | VY021328.D     | 02/26/2025       | 13:54            |
| BP-VPB-192-GW-860-862 | Q1401-08         | VY021329.D     | 02/26/2025       | 14:17            |
| BP-VPB-192-GW-880-882 | Q1401-09         | VY021330.D     | 02/26/2025       | 14:40            |
| VSTDCCC050EC          | VSTDCCC050       | VY021344.D     | 02/26/2025       | 20:06            |