

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM

SAS No.: Q1402

SDG NO.: Q1402

Lab File ID: BN036408.D

DFTPP Injection Date: 02/10/2025

Instrument ID: BNA_N

DFTPP Injection Time: 11:46

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	51.4
68	Less than 2.0% of mass 69	0.3 (0.7) 1
69	Mass 69 relative abundance	47.7
70	Less than 2.0% of mass 69	0.3 (0.6) 1
127	10.0 - 80.0% of mass 198	48.3
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7
275	10.0 - 60.0% of mass 198	24.7
365	Greater than 1% of mass 198	3.3
441	Present, but less than mass 443	7.1
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	10.5 (20.1) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC0.1	SSTDICC0.1	BN036409.D	02/10/2025	12:25
SSTDICC0.2	SSTDICC0.2	BN036410.D	02/10/2025	13:01
SSTDICCC0.4	SSTDICCC0.4	BN036411.D	02/10/2025	13:36
SSTDICC0.8	SSTDICC0.8	BN036412.D	02/10/2025	14:12
SSTDICC1.6	SSTDICC1.6	BN036413.D	02/10/2025	14:48
SSTDICC3.2	SSTDICC3.2	BN036414.D	02/10/2025	15:24
SSTDICC5.0	SSTDICC5.0	BN036415.D	02/10/2025	16:00