

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1403 SAS No.: Q1403 SDG NO.: Q1403

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/26/2025 02/26/2025

Client Sample No.: CCAL01 Date Analyzed: 02/26/2025

Lab Sample No.: AR1660CCC400 Data File : PQ070011.D Time Analyzed: 15:51

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.485	4.385	4.585	265191200.000	272527000.000	2.8
Aroclor-1016-2	4.504	4.403	4.603	455488600.000	440471000.000	-3.3
Aroclor-1016-3	4.561	4.462	4.662	299813200.000	286973000.000	-4.3
Aroclor-1016-4	4.651	4.551	4.751	237375700.000	232464000.000	-2.1
Aroclor-1016-5	4.935	4.835	5.035	232879600.000	228914000.000	-1.7
Aroclor-1260-1	6.025	5.926	6.126	397928000.000	387817000.000	-2.5
Aroclor-1260-2	6.281	6.182	6.382	429095700.000	423510000.000	-1.3
Aroclor-1260-3	6.632	6.532	6.732	347609000.000	342531000.000	-1.5
Aroclor-1260-4	6.846	6.746	6.946	376519100.000	371117000.000	-1.4
Aroclor-1260-5	7.152	7.052	7.252	662059400.000	663860000.000	0.3
DCB	8.504	8.404	8.604	4467595000.000	4409720000.000	-1.3
TCX	3.405	3.304	3.504	10442500000.000	10209900000.000	-2.3

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1403 SAS No.: Q1403 SDG NO.: Q1403

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/26/2025 02/26/2025

Client Sample No.: CCAL01 Date Analyzed: 02/26/2025

Lab Sample No.: AR1660CCC400 Data File : PQ070011.D Time Analyzed: 15:51

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.787	3.687	3.887	217195800.000	219495000.000	1.1
Aroclor-1016-2	3.802	3.702	3.902	408040000.000	343379000.000	-15.8
Aroclor-1016-3	3.963	3.863	4.063	204902400.000	193557000.000	-5.5
Aroclor-1016-4	4.011	3.911	4.111	189779300.000	172713000.000	-9.0
Aroclor-1016-5	4.206	4.106	4.306	220863400.000	205447000.000	-7.0
Aroclor-1260-1	5.201	5.100	5.300	398528300.000	374300000.000	-6.1
Aroclor-1260-2	5.390	5.290	5.490	461970500.000	441007000.000	-4.5
Aroclor-1260-3	5.532	5.431	5.631	439916700.000	417353000.000	-5.1
Aroclor-1260-4	5.993	5.893	6.093	390238500.000	365461000.000	-6.4
Aroclor-1260-5	6.237	6.137	6.337	814273700.000	793284000.000	-2.6
DCB	7.550	7.450	7.650	5951967000.000	5599390000.000	-5.9
TCX	2.784	2.684	2.884	8098077000.000	7527800000.000	-7.1