

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1403 SAS No.: Q1403 SDG NO.: Q1403

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/26/2025 02/26/2025

Client Sample No.: CCAL02 Date Analyzed: 02/26/2025

Lab Sample No.: AR1660CCC400 Data File : PQ070015.D Time Analyzed: 16:50

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.485	4.385	4.585	265191200.000	272691000.000	2.8
Aroclor-1016-2	4.503	4.403	4.603	455488600.000	439956000.000	-3.4
Aroclor-1016-3	4.561	4.462	4.662	299813200.000	284814000.000	-5.0
Aroclor-1016-4	4.650	4.551	4.751	237375700.000	232122000.000	-2.2
Aroclor-1016-5	4.935	4.835	5.035	232879600.000	229971000.000	-1.3
Aroclor-1260-1	6.024	5.926	6.126	397928000.000	387539000.000	-2.6
Aroclor-1260-2	6.280	6.182	6.382	429095700.000	432665000.000	0.8
Aroclor-1260-3	6.632	6.532	6.732	347609000.000	340026000.000	-2.2
Aroclor-1260-4	6.845	6.746	6.946	376519100.000	371330000.000	-1.4
Aroclor-1260-5	7.151	7.052	7.252	662059400.000	674763000.000	1.9
DCB	8.504	8.404	8.604	4467595000.000	4461590000.000	-0.1
TCX	3.404	3.304	3.504	10442500000.000	10103000000.000	-3.3

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1403 SAS No.: Q1403 SDG NO.: Q1403

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/26/2025 02/26/2025

Client Sample No.: CCAL02 Date Analyzed: 02/26/2025

Lab Sample No.: AR1660CCC400 Data File : PQ070015.D Time Analyzed: 16:50

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	3.786	3.687	3.887	217195800.000	219189000.000	0.9
Aroclor-1016-2	3.802	3.702	3.902	408040000.000	341350000.000	-16.3
Aroclor-1016-3	3.962	3.863	4.063	204902400.000	190788000.000	-6.9
Aroclor-1016-4	4.011	3.911	4.111	189779300.000	169318000.000	-10.8
Aroclor-1016-5	4.206	4.106	4.306	220863400.000	200366000.000	-9.3
Aroclor-1260-1	5.201	5.100	5.300	398528300.000	365807000.000	-8.2
Aroclor-1260-2	5.389	5.290	5.490	461970500.000	437649000.000	-5.3
Aroclor-1260-3	5.531	5.431	5.631	439916700.000	412053000.000	-6.3
Aroclor-1260-4	5.992	5.893	6.093	390238500.000	366523000.000	-6.1
Aroclor-1260-5	6.236	6.137	6.337	814273700.000	800579000.000	-1.7
DCB	7.550	7.450	7.650	5951967000.000	5641450000.000	-5.2
TCX	2.784	2.684	2.884	8098077000.000	7347340000.000	-9.3