

Surrogate Summary

SDG No.: Q1403

Client: Chemtech Consulting Group

Analytical Method: SFAM_PCB

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PQ069997.D	PIBLK-PQ069997.D	TCX	1	20	0.20	102		30	150
		DCB	1	40	0.42	106		30	150
		TCX	2	20	0.21	105		30	150
		DCB	2	40	0.41	102		30	150
I.BLK-PQ070010.D	PIBLK-PQ070010.D	TCX	1	20	0.21	104		30	150
		DCB	1	40	0.43	109		30	150
		TCX	2	20	0.21	105		30	150
		DCB	2	40	0.42	104		30	150
Q1403-13	PCB-GPC2-BLANK	TCX	1	20	0	0	*	30	150
		DCB	1	40	0	0	*	30	150
		TCX	2	20	0	0	*	30	150
		DCB	2	40	0	0	*	30	150
Q1403-14	PCB-GPC2-BLANK-SPIKE	TCX	1	20	0	0	*	30	150
		DCB	1	40	0	0	*	30	150
		TCX	2	20	0	0	*	30	150
		DCB	2	40	0	0	*	30	150
I.BLK-PQ070014.D	PIBLK-PQ070014.D	TCX	1	20	0.21	103		30	150
		DCB	1	40	0.43	107		30	150
		TCX	2	20	0.20	102		30	150
		DCB	2	40	0.42	104		30	150