

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1403 SAS No.: Q1403 SDG NO.: Q1403

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/18/2025 02/18/2025

Client Sample No.: CCAL01 Date Analyzed: 03/03/2025

Lab Sample No.: INDA318 Data File : PD087939.D Time Analyzed: 15:51

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
4,4'-DDD	6.708	6.638	6.778	2568240.000	2257750.000	-12.1
4,4'-DDT	7.025	6.955	7.095	2665660.000	2348810.000	-11.9
alpha-BHC	4.002	3.952	4.052	4131350.000	3549620.000	-14.1
DCB	9.079	8.979	9.179	3294960.000	2949140.000	-10.5
Dieldrin	6.351	6.281	6.421	3644250.000	3257670.000	-10.6
Endosulfan I	6.078	6.008	6.148	3360700.000	2961680.000	-11.9
Endrin	6.578	6.508	6.648	3028440.000	2685380.000	-11.3
gamma-BHC (Lindane)	4.333	4.283	4.383	4101440.000	3510460.000	-14.4
Heptachlor	4.933	4.883	4.983	4025040.000	3528480.000	-12.3
Methoxychlor	7.497	7.427	7.567	1464130.000	1311540.000	-10.4
TCX	3.553	3.503	3.603	2086660.000	1760820.000	-15.6

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1403 SAS No.: Q1403 SDG NO.: Q1403

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/18/2025 02/18/2025

Client Sample No.: CCAL01 Date Analyzed: 03/03/2025

Lab Sample No.: INDA318 Data File : PD087939.D Time Analyzed: 15:51

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.934	5.864	6.004	14429100.000	14262200.000	-1.2
4,4'-DDT	6.189	6.119	6.259	14643200.000	15008800.000	2.5
alpha-BHC	3.397	3.347	3.447	21057100.000	21319200.000	1.2
DCB	8.079	7.979	8.179	15809900.000	15636700.000	-1.1
Dieldrin	5.518	5.448	5.588	17907700.000	17982300.000	0.4
Endosulfan I	5.252	5.182	5.322	16918600.000	17017000.000	0.6
Endrin	5.794	5.724	5.864	15984600.000	15944300.000	-0.3
gamma-BHC (Lindane)	3.733	3.683	3.783	19255900.000	19680900.000	2.2
Heptachlor	4.087	4.037	4.137	19870200.000	19946500.000	0.4
Methoxychlor	6.759	6.689	6.829	6618460.000	6314390.000	-4.6
TCX	2.884	2.834	2.934	13450900.000	13572800.000	0.9