

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: CHEM02
Lab Code: CHEM Case No.: Q1403 SAS No.: Q1403 SDG NO.: Q1403
Lab File ID: VX045067.D BFB Injection Date: 02/28/2025
Instrument ID: MSVOA_X BFB Injection Time: 01:03
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	20.7
75	30.0 - 60.0% of mass 95	53.6
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.6
173	Less than 2.0% of mass 174	0.7 (1) 1
174	50.0 - 100.0% of mass 95	73.8
175	5.0 - 9.0% of mass 174	5.8 (7.9) 1
176	95.0 - 101.0% of mass 174	70.6 (95.6) 1
177	5.0 - 9.0% of mass 176	4.3 (6.2) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC001	VSTDIC001	VX045068.D	02/28/2025	01:27
VSTDIC005	VSTDIC005	VX045069.D	02/28/2025	02:13
VSTDIC020	VSTDIC020	VX045070.D	02/28/2025	02:37
VSTDIC050	VSTDIC050	VX045071.D	02/28/2025	03:00
VSTDIC100	VSTDIC100	VX045072.D	02/28/2025	03:23
VSTDIC150	VSTDIC150	VX045073.D	02/28/2025	03:47