

Laboratory Control Sample Summary

Client: Tully Environmental, Inc

SDG No.: Q1404

Project: Transfer Station-SPDES

Run No.: LB134784

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134784BS							
TSS	mg/L	550	532		97	1	90-110	02/21/2025