

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TECH05

Lab Code: CHEM Case No.: Q1405 SAS No.: Q1405 SDG No.: Q1405

Instrument ID: MSVOA_N Calibration Date/Time: 02/24/2025 11:37

Lab File ID: VN085829.D Init. Calib. Date(s): 02/18/2025 02/18/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:09 14:18

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.680	0.700		2.94	20
Chloromethane	0.682	0.660	0.1	-3.23	20
Vinyl Chloride	0.723	0.719		-0.55	20
Bromomethane	0.462	0.437		-5.41	20
Chloroethane	0.479	0.461		-3.76	20
Trichlorofluoromethane	1.046	1.056		0.96	20
1,1,2-Trichlorotrifluoroethane	0.605	0.643		6.28	20
1,1-Dichloroethene	0.548	0.569		3.83	20
Acetone	0.205	0.217		5.85	20
Carbon Disulfide	1.625	1.541		-5.17	20
Methyl tert-butyl Ether	1.629	1.821		11.73	20
Methyl Acetate	0.705	0.776		10.07	20
Methylene Chloride	0.660	0.670		1.51	20
trans-1,2-Dichloroethene	0.590	0.607		2.88	20
1,1-Dichloroethane	1.106	1.163	0.1	5.15	20
Cyclohexane	0.926	0.906		-2.16	20
2-Butanone	0.302	0.356		17.88	20
Carbon Tetrachloride	0.550	0.590		7.27	20
cis-1,2-Dichloroethene	0.685	0.716		4.53	20
Bromochloromethane	0.454	0.496		9.25	20
Chloroform	1.156	1.205		4.24	20
1,1,1-Trichloroethane	1.028	1.063		3.4	20
Methylcyclohexane	0.455	0.502		10.33	20
Benzene	1.484	1.612		8.63	20
1,2-Dichloroethane	0.483	0.508		5.18	20
Trichloroethene	0.364	0.369		1.37	20
1,2-Dichloropropane	0.358	0.395		10.34	20
Bromodichloromethane	0.541	0.579		7.02	20
4-Methyl-2-Pentanone	0.367	0.466		26.98	20
Toluene	0.870	1.009		15.98	20
t-1,3-Dichloropropene	0.500	0.566		13.2	20
cis-1,3-Dichloropropene	0.548	0.613		11.86	20
1,1,2-Trichloroethane	0.346	0.386		11.56	20
2-Hexanone	0.253	0.338		33.6	20
Dibromochloromethane	0.403	0.459		13.9	20
1,2-Dibromoethane	0.327	0.376		14.98	20
Tetrachloroethene	0.387	0.388		0.26	20
Chlorobenzene	1.145	1.187	0.3	3.67	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TECH05

Lab Code: CHEM Case No.: Q1405 SAS No.: Q1405 SDG No.: Q1405

Instrument ID: MSVOA_N Calibration Date/Time: 02/24/2025 11:37

Lab File ID: VN085829.D Init. Calib. Date(s): 02/18/2025 02/18/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:09 14:18

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.794	2.058		14.72	20
m/p-Xylenes	0.682	0.815		19.5	20
o-Xylene	0.648	0.761		17.44	20
Styrene	1.059	1.326		25.21	20
Bromoform	0.306	0.351	0.1	14.71	20
Isopropylbenzene	3.425	3.702		8.09	20
1,1,2,2-Tetrachloroethane	1.179	1.192	0.3	1.1	20
1,3-Dichlorobenzene	1.734	1.758		1.38	20
1,4-Dichlorobenzene	1.777	1.724		-2.98	20
1,2-Dichlorobenzene	1.667	1.700		1.92	20
1,2-Dibromo-3-Chloropropane	0.216	0.219		1.39	20
1,2,4-Trichlorobenzene	0.794	0.808		1.76	20
1,2,3-Trichlorobenzene	0.798	0.802		0.5	20
1,2-Dichloroethane-d4	0.648	0.721		11.27	20
Dibromofluoromethane	0.327	0.384		17.43	20
Toluene-d8	1.190	1.428		20	20
4-Bromofluorobenzene	0.392	0.497		26.79	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.