

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TECH05

Lab Code: CHEM Case No.: Q1405 SAS No.: Q1405 SDG No.: Q1405

Instrument ID: MSVOA_N Calibration Date/Time: 02/24/2025 11:37

Lab File ID: VN085829.D Init. Calib. Date(s): 02/18/2025 02/18/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:09 14:18

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Styrene	50.000	54.50	-8.991997	20