

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TECH05
 Lab Code: CHEM Case No.: Q1405 SAS No.: Q1405 SDG NO.: Q1405
 Lab File ID: VN085802.D Date Analyzed: 02/21/2025
 Instrument ID: MSVOA_N Time Analyzed: 14:43
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	205820	13.788				
UPPER LIMIT	411640	14.288				
LOWER LIMIT	102910	13.288				
EPA SAMPLE NO.						
MW-01-021325	159960	13.79				
MW-04-021325	183668	13.79				
MW-23-021425	167368	13.79				
VN0221WBL01	150905	13.79				
VN0221WBS01	206616	13.79				
VN0221WBSD01	190177	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.