

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TECH05
 Lab Code: CHEM Case No.: Q1405 SAS No.: Q1405 SDG NO.: Q1405
 Lab File ID: VN085829.D Date Analyzed: 02/24/2025
 Instrument ID: MSVOA_N Time Analyzed: 11:37
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	208250	13.788				
UPPER LIMIT	416500	14.288				
LOWER LIMIT	104125	13.288				
EPA SAMPLE NO.						
MW-04-021325DL	127235	13.79				
MW-20D-021325	144563	13.79				
MW-20D-FD-021325	144264	13.79				
MW-22-021925	123885	13.79				
MW-26-021825	143476	13.79				
MOS-11R-021325	143066	13.79				
MOS-13-021925	124231	13.79				
MW-30S-021825	131978	13.79				
TB-02-021325	137032	13.79				
VN0224WBL01	141296	13.79				
VN0224WBS01	189869	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.