

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: TECH05
Lab Code: CHEM Case No.: Q1405 SAS No.: Q1405 SDG NO.: Q1405
Lab File ID: VN085771.D BFB Injection Date: 02/18/2025
Instrument ID: MSVOA_N BFB Injection Time: 10:35
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	18.1
75	30.0 - 60.0% of mass 95	49.7
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.5
173	Less than 2.0% of mass 174	1.4 (1.8) 1
174	50.0 - 100.0% of mass 95	81.7
175	5.0 - 9.0% of mass 174	6 (7.4) 1
176	95.0 - 101.0% of mass 174	78.9 (96.5) 1
177	5.0 - 9.0% of mass 176	5.4 (6.9) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDICC100	VSTDICC100	VN085772.D	02/18/2025	11:09
VSTDICCC050	VSTDICCC050	VN085773.D	02/18/2025	11:32
VSTDICC010	VSTDICC010	VN085775.D	02/18/2025	12:20
VSTDICC005	VSTDICC005	VN085776.D	02/18/2025	12:43
VSTDICC001	VSTDICC001	VN085777.D	02/18/2025	13:07
VSTDICC020	VSTDICC020	VN085779.D	02/18/2025	14:18