

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Homeowner **SDG No.:** Q1410
Contract: HOME01 **Lab Code:** CHEM **Case No.:** Q1410 **SAS No.:** Q1410
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Copper	531	510	104	95 - 105	P	03/06/2025	10:50	LB134928
	Lead	1010	1000	100	95 - 105	P	03/06/2025	10:50	LB134928

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Homeowner **SDG No.:** Q1410
Contract: HOME01 **Lab Code:** CHEM **Case No.:** Q1410 **SAS No.:** Q1410
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Copper	22.8	20.0	114	80 - 120	P	03/06/2025	11:38	LB134928
	Lead	11.9	12.0	99	80 - 120	P	03/06/2025	11:38	LB134928

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Homeowner SDG No.: Q1410
Contract: HOME01 Lab Code: CHEM Case No.: Q1410 SAS No.: Q1410
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Copper	1290	1250	103	90 - 110	P	03/06/2025	12:19	LB134928
	Lead	5010	5000	100	90 - 110	P	03/06/2025	12:19	LB134928
CCV02	Copper	1250	1250	100	90 - 110	P	03/06/2025	13:10	LB134928
	Lead	4830	5000	97	90 - 110	P	03/06/2025	13:10	LB134928
CCV03	Copper	1270	1250	101	90 - 110	P	03/06/2025	14:00	LB134928
	Lead	4830	5000	97	90 - 110	P	03/06/2025	14:00	LB134928
CCV04	Copper	1250	1250	100	90 - 110	P	03/06/2025	14:50	LB134928
	Lead	4910	5000	98	90 - 110	P	03/06/2025	14:50	LB134928
CCV05	Copper	1250	1250	100	90 - 110	P	03/06/2025	15:53	LB134928
	Lead	4940	5000	99	90 - 110	P	03/06/2025	15:53	LB134928