



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Homeowner **SDG No.:** Q1410
Contract: HOME01 **Lab Code:** CHEM **Case No.:** Q1410 **SAS No.:** Q1410
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Copper	22.6	20.0	113	40 - 160	P	03/06/2025	11:47	LB134928
	Lead	12.6	12.0	105	40 - 160	P	03/06/2025	11:47	LB134928